

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
 Data File : VY016208.D
 Acq On : 03 Nov 2023 12:26
 Operator : SY/MD
 Sample : 05206-02RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOCRE

Quant Time: Nov 03 23:01:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	49333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	84258	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	73869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	30652	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	26584	57.588	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.180%
35) Dibromofluoromethane	7.734	113	24331	48.437	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.880%
50) Toluene-d8	10.191	98	94468	47.433	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.860%
62) 4-Bromofluorobenzene	12.489	95	29525	43.354	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.700%
Target Compounds						
20) Methylene Chloride	4.729	84	4298	6.505	ug/l #	84
52) Toluene	10.252	92	1793	1.329	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
Data File : VY016208.D
Acq On : 03 Nov 2023 12:26
Operator : SY/MD
Sample : 05206-02RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOCRE

Quant Time: Nov 03 23:01:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 01 03:33:29 2023
Response via : Initial Calibration

