

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
 Data File : VY016212.D
 Acq On : 03 Nov 2023 14:00
 Operator : SY/MD
 Sample : 05229-01
 Misc : 3.61g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-1-110223

Quant Time: Nov 03 23:02:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	113060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	198508	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	190751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	90237	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	65907	62.298	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.600%
35) Dibromofluoromethane	7.728	113	56632	47.853	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.700%
50) Toluene-d8	10.191	98	233414	49.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.489	95	83897	52.290	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.580%
Target Compounds						
20) Methylene Chloride	4.728	84	16175	12.764	ug/l	92
84) 1,2,4-Trimethylbenzene	13.129	105	7202	1.327	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
Data File : VY016212.D
Acq On : 03 Nov 2023 14:00
Operator : SY/MD
Sample : 05229-01
Misc : 3.61g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-1-110223

Quant Time: Nov 03 23:02:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 01 03:33:29 2023
Response via : Initial Calibration

