

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
 Data File : VY016214.D
 Acq On : 03 Nov 2023 15:23
 Operator : SY/MD
 Sample : BLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLK

Quant Time: Nov 03 23:03:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	117123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	204660	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	192899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	92845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	63583	58.016	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.040%
35) Dibromofluoromethane	7.728	113	56708	46.477	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.960%
50) Toluene-d8	10.191	98	241083	49.836	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.489	95	91055	55.045	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.100%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.722	84	14576	10.681	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
Data File : VY016214.D
Acq On : 03 Nov 2023 15:23
Operator : SY/MD
Sample : BLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLK

Quant Time: Nov 03 23:03:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 01 03:33:29 2023
Response via : Initial Calibration

