

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
 Data File : VY016218.D
 Acq On : 03 Nov 2023 18:25
 Operator : SY/MD
 Sample : VY1103SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1103SBL02

Quant Time: Nov 03 23:05:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	156310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	262248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	229087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	103784	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	84663	57.884	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.760%
35) Dibromofluoromethane	7.734	113	79950	51.137	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.280%
50) Toluene-d8	10.191	98	316314	51.029	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.060%
62) 4-Bromofluorobenzene	12.489	95	104272	49.193	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110323\
Data File : VY016218.D
Acq On : 03 Nov 2023 18:25
Operator : SY/MD
Sample : VY1103SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1103SBL02

Quant Time: Nov 03 23:05:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 01 03:33:29 2023
Response via : Initial Calibration

