

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110421\
 Data File : VY006626.D
 Acq On : 04 Nov 2021 17:51
 Operator : SY/MD
 Sample : M4448-02
 Misc : 4.17G/5ML/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-224

Quant Time: Nov 05 14:00:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	131055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	225574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	211045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	84303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	96906	60.089	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 120.180%		
35) Dibromofluoromethane	7.728	113	70643	56.187	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 112.380%		
50) Toluene-d8	10.185	98	288886	54.202	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 108.400%		
62) 4-Bromofluorobenzene	12.483	95	90759	50.551	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 101.100%		
Target Compounds						
20) Methylene Chloride	4.735	84	21141	11.569	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110421\
Data File : VY006626.D
Acq On : 04 Nov 2021 17:51
Operator : SY/MD
Sample : M4448-02
Misc : 4.17G/5ML/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-224

Quant Time: Nov 05 14:00:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

