

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110421\
 Data File : VY006634.D
 Acq On : 04 Nov 2021 21:00
 Operator : SY/MD
 Sample : M4485-05
 Misc : 6.03G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AG-6

Quant Time: Nov 05 14:03:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	128057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	217922	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	201963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	80248	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	101597	64.472	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.940%	
35) Dibromofluoromethane	7.728	113	68443	56.349	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.700%	
50) Toluene-d8	10.185	98	281202	54.613	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.220%	
62) 4-Bromofluorobenzene	12.483	95	90352	52.091	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.180%	
Target Compounds						
16) Acetone	3.954	43	5791	12.762	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110421\
Data File : VY006634.D
Acq On : 04 Nov 2021 21:00
Operator : SY/MD
Sample : M4485-05
Misc : 6.03G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AG-6

Quant Time: Nov 05 14:03:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

