

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110524\
 Data File : VY020164.D
 Acq On : 05 Nov 2024 14:08
 Operator : SY/MD
 Sample : P4693-03
 Misc : 3.54g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-G5-WC-VOC

Quant Time: Nov 06 00:25:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	157006	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	329019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	308956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	116313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	115100	58.738	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.480%
35) Dibromofluoromethane	7.640	113	109102	52.155	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.300%
50) Toluene-d8	10.109	98	411168	49.379	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.760%
62) 4-Bromofluorobenzene	12.407	95	129603	47.919	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	95.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110524\
Data File : VY020164.D
Acq On : 05 Nov 2024 14:08
Operator : SY/MD
Sample : P4693-03
Misc : 3.54g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-G5-WC-VOC

Quant Time: Nov 06 00:25:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

