

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110624\
 Data File : VY020187.D
 Acq On : 06 Nov 2024 17:27
 Operator : SY/MD
 Sample : P4675-05
 Misc : 6.65g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-5

Quant Time: Nov 07 00:29:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	1297	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1970	50.000	ug/l	# 0.00
63) Chlorobenzene-d5	11.420	117	1013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	134	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	1066	65.854	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.700%
35) Dibromofluoromethane	7.640	113	406	32.415	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	64.820%
50) Toluene-d8	10.109	98	1879	37.688	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	75.380%
62) 4-Bromofluorobenzene	12.414	95	258	15.932	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	31.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110624\
Data File : VY020187.D
Acq On : 06 Nov 2024 17:27
Operator : SY/MD
Sample : P4675-05
Misc : 6.65g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-5

Quant Time: Nov 07 00:29:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

