

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110624\
 Data File : VY020190.D
 Acq On : 06 Nov 2024 18:37
 Operator : SY/MD
 Sample : P4706-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-110424

Quant Time: Nov 07 00:30:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	10098	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	17828	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	12620	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	2329	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	6387	50.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.360%
35) Dibromofluoromethane	7.634	113	6155	54.301	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.600%
50) Toluene-d8	10.103	98	20201	44.773	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.540%
62) 4-Bromofluorobenzene	12.414	95	4041	27.574	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	55.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110624\
Data File : VY020190.D
Acq On : 06 Nov 2024 18:37
Operator : SY/MD
Sample : P4706-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-110424

Quant Time: Nov 07 00:30:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

