

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110723\  
 Data File : VY016248.D  
 Acq On : 07 Nov 2023 12:45  
 Operator : SY/MD  
 Sample : BLK  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 BLK

Quant Time: Nov 07 23:44:32 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y103123S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Nov 01 03:33:29 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	7.795	168	108747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	191464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	183413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	83818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	61138	60.082	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.160%
35) Dibromofluoromethane	7.722	113	57186	50.099	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.200%
50) Toluene-d8	10.185	98	225812	49.896	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.800%
62) 4-Bromofluorobenzene	12.489	95	77779	50.260	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.520%

Target Compounds Qvalue

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY110723\  
 Data File : VY016248.D  
 Acq On : 07 Nov 2023 12:45  
 Operator : SY/MD  
 Sample : BLK  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 BLK

Quant Time: Nov 07 23:44:32 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y103123S.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Nov 01 03:33:29 2023  
 Response via : Initial Calibration

