

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110724\
 Data File : VY020214.D
 Acq On : 07 Nov 2024 18:04
 Operator : SY/MD
 Sample : P4739-07
 Misc : 4.32g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Nov 08 00:31:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	157154	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	340661	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	345228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	158032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	123635	63.035	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.060%	
35) Dibromofluoromethane	7.640	113	115819	53.474	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.940%	
50) Toluene-d8	10.103	98	456320	52.929	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 105.860%	
62) 4-Bromofluorobenzene	12.402	95	185323	66.179	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 132.360%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.854	59	1034	7.329	ug/l	# 88
16) Acetone	3.873	43	12104	26.536	ug/l	# 85
20) Methylene Chloride	4.623	84	10334	5.043	ug/l	94
25) 2-Butanone	6.909	43	5173	8.366	ug/l	# 87
31) Cyclohexane	7.707	56	6631	1.863	ug/l	# 54
36) 1,1-Dichloropropene	7.707	75	16072	4.804	ug/l	# 49
38) Carbon Tetrachloride	7.707	117	17830	5.299	ug/l	# 15
39) Methylcyclohexane	9.110	83	25740	6.062	ug/l	# 86
40) Benzene	8.079	78	18661	1.911	ug/l	99
43) Isopropyl Acetate	8.183	43	3509	1.144	ug/l	# 47
48) Methyl methacrylate	9.164	41	6301	4.435	ug/l	# 38
51) 4-Methyl-2-Pentanone	10.097	43	10682	6.758	ug/l	98
55) 1,1,2-Trichloroethane	10.542	97	61720	39.209	ug/l	# 21
56) Ethyl methacrylate	10.445	69	34337	15.207	ug/l	# 1
59) 2-Hexanone	10.731	43	224189	197.845	ug/l	# 23
73) Isopropylbenzene	12.255	105	116619	9.238	ug/l	97
74) N-amyl acetate	12.048	43	395425	125.303	ug/l	# 52
75) 1,1,2,2-Tetrachloroethane	12.481	83	37598	17.029	ug/l	# 73
76) 1,2,3-Trichloropropane	12.402	75	93588	63.176	ug/l	# 100
78) n-propylbenzene	12.591	91	142443	9.066	ug/l	94
80) 1,3,5-Trimethylbenzene	12.895	105	27340	2.698	ug/l	71
81) trans-1,4-Dichloro-2-b...	12.402	75	93588	135.889	ug/l	# 24
82) 4-Chlorotoluene	12.725	91	13119	1.475	ug/l	# 59
83) tert-Butylbenzene	12.993	119	67337	7.598	ug/l	84
84) 1,2,4-Trimethylbenzene	12.895	105	27340	2.723	ug/l	75
85) sec-Butylbenzene	13.176	105	210791	15.493	ug/l	78
89) n-Butylbenzene	13.615	91	110291	10.077	ug/l	94
90) Hexachloroethane	13.895	117	29267	13.554	ug/l	# 12
92) 1,2-Dibromo-3-Chloropr...	14.298	75	1401	4.184	ug/l	# 4
95) Naphthalene	15.115	128	22023	4.865	ug/l	# 1
96) 1,2,3-Trichlorobenzene	15.352	180	9722	4.729	ug/l	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 08 00:31:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 157154
Ion Ratio Lower Upper
168 100
99 63.8 39.1 58.7#



#11
Tert butyl alcohol
Concen: 7.329 ug/l
RT: 4.854 min Scan# 514
Delta R.T. -0.006 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 59 Resp: 1034
Ion Ratio Lower Upper
59 100
57 0.0 3.0 4.6#





#16

Acetone

Concen: 26.536 ug/l

RT: 3.873 min Scan# 353

Delta R.T. 0.000 min


Lab File: VY020214.D

Acq: 07 Nov 2024 18:04

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion: 43 Resp: 12104

Ion Ratio Lower Upper

43 100

58 30.5 31.7 47.5#



#20

Methylene Chloride

Concen: 5.043 ug/l

RT: 4.623 min Scan# 476

Delta R.T. 0.007 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04

Tgt Ion: 84 Resp: 10334




Ion Ratio Lower Upper


84 100

49 112.7 84.4 126.6

51 36.4 26.0 39.0

86 64.5 49.5 74.3









#25
2-Butanone
Concen: 8.366 ug/l
RT: 6.909 min Scan# 81
Delta R.T. 0.019 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 5173
Ion Ratio Lower Upper
43 100
72 26.7 27.3 40.9#



#31
Cyclohexane
Concen: 1.863 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.007 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 56 Resp: 6631
Ion Ratio Lower Upper
56 100
69 120.1 28.7 43.1#
84 106.8 77.7 116.5





#33

1,2-Dichloroethane-d4

Concen: 63.035 ug/l

RT: 8.061 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 65 Resp: 123635

Ion Ratio Lower Upper

65 100

67 52.3 0.0 109.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04




Tgt Ion: 114 Resp: 340661


Ion Ratio Lower Upper

114 100

63 24.1 0.0 35.0


88 16.3 0.0 27.2








#35
Dibromofluoromethane
Concen: 53.474 ug/l
RT: 7.640 min Scan# 91
Delta R.T. 0.007 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:113 Resp: 115819
Ion Ratio Lower Upper
113 100
111 103.7 82.2 123.4
192 15.4 15.9 23.9#



#36
1,1-Dichloropropene
Concen: 4.804 ug/l
RT: 7.707 min Scan# 982
Delta R.T. -0.128 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04


Tgt Ion: 75 Resp: 16072
Ion Ratio Lower Upper
75 100
110 10.3 19.0 57.0#
77 0.0 25.3 37.9#








#38
Carbon Tetrachloride
Concen: 5.299 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.109 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:117 Resp: 17830
Ion Ratio Lower Upper
117 100
119 3.4 75.8 113.6#
121 0.0 24.0 36.0#



#39
Methylcyclohexane
Concen: 6.062 ug/l
RT: 9.110 min Scan# 1212
Delta R.T. 0.000 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 83 Resp: 25740
Ion Ratio Lower Upper
83 100
55 82.9 51.5 77.3#
98 48.6 40.5 60.7









#40
Benzene
Concen: 1.911 ug/l
RT: 8.079 min Scan# 1043
Delta R.T. 0.000 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 78 Resp: 18661
Ion Ratio Lower Upper
78 100
77 24.9 19.5 29.3



#43
Isopropyl Acetate
Concen: 1.144 ug/l
RT: 8.183 min Scan# 1060
Delta R.T. -0.012 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 43 Resp: 3509
Ion Ratio Lower Upper
43 100
61 0.0 27.9 41.9#
87 0.0 13.1 19.7#





#48

Methyl methacrylate

Concen: 4.435 ug/l

RT: 9.164 min Scan# 1

Delta R.T. -0.054 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 41 Resp: 6301

Ion Ratio Lower Upper

41 100

69 12.6 84.1 126.1#

39 48.8 43.6 65.4



#50

Toluene-d8

Concen: 52.929 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.006 min

Lab File: VY020214.D




Acq: 07 Nov 2024 18:04


Tgt Ion: 98 Resp: 456320

Ion Ratio Lower Upper

98 100

100 63.0 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 6.758 ug/l
RT: 10.097 min Scan# 11
Delta R.T. 0.098 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 10682
Ion Ratio Lower Upper
43 100
58 43.1 35.7 53.5



#55
1,1,2-Trichloroethane
Concen: 39.209 ug/l
RT: 10.542 min Scan# 1447
Delta R.T. -0.030 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 97 Resp: 61720
Ion Ratio Lower Upper
97 100
83 8.6 67.5 101.3#
85 2.0 42.9 64.3#
99 0.9 46.6 70.0#





#56

Ethyl methacrylate

Concen: 15.207 ug/l

RT: 10.445 min Scan# 1430

Delta R.T. 0.007 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 69 Resp: 34337

Ion Ratio Lower Upper

69 100

41 177.6 42.3 63.5#

39 92.9 22.9 34.3#



#59

2-Hexanone

Concen: 197.845 ug/l

RT: 10.731 min Scan# 1478

Delta R.T. -0.030 min

Lab File: VY020214.D




Acq: 07 Nov 2024 18:04


Tgt Ion: 43 Resp: 224189

Ion Ratio Lower Upper

43 100

58 2.9 31.4 94.0#









#62
4-Bromofluorobenzene
Concen: 66.179 ug/l
RT: 12.402 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 95 Resp: 185323
Ion Ratio Lower Upper
95 100
174 63.4 0.0 175.6
176 61.3 0.0 171.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.006 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 117 Resp: 345228
Ion Ratio Lower Upper
117 100
82 54.9 42.4 63.6
119 31.4 25.9 38.9





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.347 min Scan# 1907

Delta R.T. 0.000 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:152 Resp: 158032

Ion Ratio Lower Upper

152 100

115 58.4 28.2 84.7

150 145.8 0.0 345.6



#73

Isopropylbenzene

Concen: 9.238 ug/l

RT: 12.255 min Scan# 1728

Delta R.T. 0.000 min

Lab File: VY020214.D




Acq: 07 Nov 2024 18:04


Tgt Ion:105 Resp: 116619

Ion Ratio Lower Upper

105 100

120 25.2 13.4 40.1





#74

N-amy1 acetate

Concen: 125.303 ug/l

RT: 12.048 min Scan# 1

Delta R.T. -0.024 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04

Instrument :

MSVOA_Y

ClientSampleId :

Tgt Ion: 43 Resp: 395425





Ion Ratio Lower Upper

43 100

70 8.7 41.4 62.2#

55 18.4 23.3 34.9#

61 0.0 22.6 33.8#



#75

1,1,2,2-Tetrachloroethane

Concen: 17.029 ug/l

RT: 12.481 min Scan# 1765

Delta R.T. -0.030 min

Lab File: VY020214.D

Acq: 07 Nov 2024 18:04




Tgt Ion: 83 Resp: 37598


Ion Ratio Lower Upper

83 100

131 0.0 5.0 14.9#

85 42.4 32.0 96.0









#76
1,2,3-Trichloropropane
Concen: 63.176 ug/l
RT: 12.402 min Scan# 1752
Delta R.T. -0.158 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 93588
Ion Ratio Lower Upper
75 100
77 5.0 0.0 0.0#



#78
n-propylbenzene
Concen: 9.066 ug/l
RT: 12.591 min Scan# 1783
Delta R.T. -0.006 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 91 Resp: 142443
Ion Ratio Lower Upper
91 100
120 20.1 11.5 34.4









#80
1,3,5-Trimethylbenzene
Concen: 2.698 ug/l
RT: 12.895 min Scan# 1752
Delta R.T. 0.159 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:105 Resp: 27340
Ion Ratio Lower Upper
105 100
120 29.9 25.1 75.2



#81
trans-1,4-Dichloro-2-butene
Concen: 135.889 ug/l
RT: 12.402 min Scan# 1752
Delta R.T. 0.098 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 75 Resp: 93588
Ion Ratio Lower Upper
75 100
53 10.2 66.8 100.2#
89 0.5 35.8 53.6#









#82
4-Chlorotoluene
Concen: 1.475 ug/l
RT: 12.725 min Scan# 1805
Delta R.T. -0.054 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 13119
Ion Ratio Lower Upper
91 100
126 59.0 17.6 52.9#



#83
tert-Butylbenzene
Concen: 7.598 ug/l
RT: 12.993 min Scan# 1849
Delta R.T. -0.006 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 119 Resp: 67337
Ion Ratio Lower Upper
119 100
91 78.6 31.3 93.8
134 24.0 13.4 40.1









#84
1,2,4-Trimethylbenzene
Concen: 2.723 ug/l
RT: 12.895 min Scan# 1833
Delta R.T. -0.152 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:105 Resp: 27340
Ion Ratio Lower Upper
105 100
120 29.9 23.2 69.6



#85
sec-Butylbenzene
Concen: 15.493 ug/l
RT: 13.176 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion:105 Resp: 210791
Ion Ratio Lower Upper
105 100
134 30.0 10.1 30.1









#89
n-Butylbenzene
Concen: 10.077 ug/l
RT: 13.615 min Scan# 1951
Delta R.T. -0.006 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 110291
Ion Ratio Lower Upper
91 100
92 49.9 26.6 79.7
134 30.3 12.9 38.7



#90
Hexachloroethane
Concen: 13.554 ug/l
RT: 13.895 min Scan# 1997
Delta R.T. 0.013 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04

Tgt Ion: 117 Resp: 29267
Ion Ratio Lower Upper
117 100
201 0.2 37.4 112.2#









#92
1,2-Dibromo-3-Chloropropane
Concen: 4.184 ug/l
RT: 14.298 min Scan# 2058
Delta R.T. 0.025 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 75 Resp: 1401
Ion Ratio Lower Upper
75 100
155 20.7 48.6 145.8#
157 0.0 62.7 188.2#



#95
Naphthalene
Concen: 4.865 ug/l
RT: 15.115 min Scan# 2197
Delta R.T. -0.030 min
Lab File: VY020214.D
Acq: 07 Nov 2024 18:04


Tgt Ion: 128 Resp: 22023
Ion Ratio Lower Upper
128 100
127 137.5 10.6 16.0#
129 68.4 8.6 13.0#





#96
 1,2,3-Trichlorobenzene
 Concen: 4.729 ug/l
 RT: 15.352 min Scan# 21
 Delta R.T. 0.025 min
 Lab File: VY020214.D
 Acq: 07 Nov 2024 18:04

Instrument :
 MSVOA_Y
 ClientSampleId :



Tgt Ion:180 Resp: 9722
 Ion Ratio Lower Upper
 180 100
 182 0.0 47.6 142.9#
 145 1391.6 15.4 46.1#

