

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110724\
 Data File : VY020201.D
 Acq On : 07 Nov 2024 13:00
 Operator : SY/MD
 Sample : P4667-07
 Misc : 5.41g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-F-5-VOC

Quant Time: Nov 07 14:52:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	150212	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	325915	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	305979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	111677	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	116485	62.134	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.260%	
35) Dibromofluoromethane	7.634	113	108368	52.297	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.600%	
50) Toluene-d8	10.103	98	412981	50.069	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.140%	
62) 4-Bromofluorobenzene	12.402	95	127543	47.607	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 95.220%	
Target Compounds						
16) Acetone	3.879	43	4016	9.211	ug/l	Qvalue 89
20) Methylene Chloride	4.623	84	3310	1.690	ug/l	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110724\
Data File : VY020201.D
Acq On : 07 Nov 2024 13:00
Operator : SY/MD
Sample : P4667-07
Misc : 5.41g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-F-5-VOC

Quant Time: Nov 07 14:52:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

