

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110724\
 Data File : VY020205.D
 Acq On : 07 Nov 2024 14:33
 Operator : SY/MD
 Sample : P4706-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-110424

Quant Time: Nov 07 15:05:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	139893	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	300044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	276117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	99781	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	101905	58.366	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.740%
35) Dibromofluoromethane	7.640	113	101163	53.030	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.060%
50) Toluene-d8	10.103	98	372949	49.115	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.220%
62) 4-Bromofluorobenzene	12.401	95	116554	47.256	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.520%
Target Compounds						
20) Methylene Chloride	4.616	84	2503	1.372	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110724\
Data File : VY020205.D
Acq On : 07 Nov 2024 14:33
Operator : SY/MD
Sample : P4706-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-110424

Quant Time: Nov 07 15:05:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

