

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110724\
 Data File : VY020218.D
 Acq On : 07 Nov 2024 19:38
 Operator : SY/MD
 Sample : P4732-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PPE-COMP

Quant Time: Nov 08 00:33:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	2748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	4959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	2990	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	465	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.067	65	2350	68.519	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	137.040%
35) Dibromofluoromethane	7.622	113	1615	51.223	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.440%
50) Toluene-d8	10.109	98	5639	44.932	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.860%
62) 4-Bromofluorobenzene	12.408	95	1048	25.709	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	51.420%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110724\
Data File : VY020218.D
Acq On : 07 Nov 2024 19:38
Operator : SY/MD
Sample : P4732-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PPE-COMP

Quant Time: Nov 08 00:33:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

