


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\VY110819\
Data File : VY000604.D
Acq On : 08 Nov 2019 09:48
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y103019S.M
Title : SW846 8260
Last Update : Wed Oct 30 13:30:53 2019



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1753

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	46573	PASS
75	95	30	60	55.1	126165	PASS
95	95	100	100	100.0	228864	PASS
96	95	5	9	6.9	15687	PASS
173	174	0.00	2	0.9	1488	PASS
174	95	50	100	74.5	170576	PASS
175	174	5	9	7.3	12502	PASS
176	174	95	101	97.3	166000	PASS
177	176	5	9	7.0	11549	PASS