

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110821\
 Data File : VY006687.D
 Acq On : 08 Nov 2021 17:54
 Operator : SY/MD
 Sample : M4540-01
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Nov 09 13:06:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	14826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	22661	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	18875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	6607	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	10572	57.947	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.900%
35) Dibromofluoromethane	7.722	113	6711	53.133	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.260%
50) Toluene-d8	10.185	98	27532	51.420	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.840%
62) 4-Bromofluorobenzene	12.483	95	6878	38.134	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.260%
Target Compounds						
20) Methylene Chloride	4.729	84	1695	6.923	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110821\
Data File : VY006687.D
Acq On : 08 Nov 2021 17:54
Operator : SY/MD
Sample : M4540-01
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Nov 09 13:06:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

