



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110821\
 Data File : VY006668.D
 Acq On : 08 Nov 2021 09:55
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y1022215.M
 Title : SW846 8260
 Last Update : Sat Oct 23 06:59:04 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.6	22355	PASS
75	95	30	60	58.6	51237	PASS
95	95	100	100	100.0	87435	PASS
96	95	5	9	6.7	5847	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	81.3	71061	PASS
175	174	5	9	8.5	6013	PASS
176	174	95	101	97.3	69144	PASS
177	176	5	9	6.8	4712	PASS