

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110823\
 Data File : VY016253.D
 Acq On : 08 Nov 2023 08:59
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
 Supervised By :Mahesh Dadoda 11/09/2023

Quant Time: Nov 09 00:27:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:22:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	167284	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	278278	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	236127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	104288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	7688	5.028	ug/l	0.00
Spiked Amount 50.000	Range 50	- 163	Recovery	=	10.060%#	
35) Dibromofluoromethane	7.728	113	7910	5.004	ug/l	0.00
Spiked Amount 50.000	Range 54	- 147	Recovery	=	10.000%#	
50) Toluene-d8	10.191	98	30918	4.823	ug/l	0.00
Spiked Amount 50.000	Range 58	- 134	Recovery	=	9.640%#	
62) 4-Bromofluorobenzene	12.495	95	10155	4.877	ug/l	0.00
Spiked Amount 50.000	Range 30	- 143	Recovery	=	9.760%#	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	9087	5.548	ug/l	97
3) Chloromethane	2.119	50	10902	5.683	ug/l	99
4) Vinyl Chloride	2.253	62	11074	5.368	ug/l	94
5) Bromomethane	2.661	94	7556	5.574	ug/l	88
6) Chloroethane	2.802	64	7051	5.357	ug/l	94
7) Trichlorofluoromethane	3.131	101	15824	5.273	ug/l	98
8) Diethyl Ether	3.533	74	4323	4.820	ug/l	91
9) 1,1,2-Trichlorotrifluo...	3.905	101	9710	5.372	ug/l	94
10) Methyl Iodide	4.106	142	9387	4.808	ug/l	97
11) Tert butyl alcohol	4.972	59	3273	27.338	ug/l #	91
12) 1,1-Dichloroethene	3.887	96	8675	5.132	ug/l	79
13) Acrolein	3.741	56	3324	23.018	ug/l	98
14) Allyl chloride	4.490	41	12105	5.046	ug/l #	94
15) Acrylonitrile	5.185	53	8697	24.433	ug/l	95
16) Acetone	3.954	43	7338	29.015	ug/l	100
17) Carbon Disulfide	4.210	76	26663	5.228	ug/l	100
18) Methyl Acetate	4.490	43	6436	4.789	ug/l	94
19) Methyl tert-butyl Ether	5.234	73	19602	4.791	ug/l	99
20) Methylene Chloride	4.728	84	13629	4.458	ug/l #	85
21) trans-1,2-Dichloroethene	5.234	96	10287	5.285	ug/l	84
22) Diisopropyl ether	6.130	45	25227	4.880	ug/l #	91
23) Vinyl Acetate	6.075	43	58116	23.200	ug/l	96
24) 1,1-Dichloroethane	6.033	63	16641	5.218	ug/l	96
25) 2-Butanone	6.996	43	11218	25.325	ug/l	95
26) 2,2-Dichloropropane	6.990	77	14602	5.030	ug/l	98
27) cis-1,2-Dichloroethene	6.996	96	10748	5.124	ug/l	92
28) Bromochloromethane	7.350	49	5931	5.025	ug/l	92
29) Tetrahydrofuran	7.356	42	6607	22.566	ug/l #	89
30) Chloroform	7.520	83	17652	5.317	ug/l	90
31) Cyclohexane	7.795	56	16940	5.628	ug/l #	77
32) 1,1,1-Trichloroethane	7.715	97	15809	5.195	ug/l	98
36) 1,1-Dichloropropene	7.929	75	13891	5.061	ug/l	97
37) Ethyl Acetate	7.094	43	5297	4.776	ug/l #	91
38) Carbon Tetrachloride	7.917	117	14219	4.991	ug/l	94
39) Methylcyclohexane	9.197	83	16915	4.945	ug/l	89
40) Benzene	8.173	78	38990	5.056	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110823\
 Data File : VY016253.D
 Acq On : 08 Nov 2023 08:59
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
 Supervised By :Mahesh Dadoda 11/09/2023

Quant Time: Nov 09 00:27:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:22:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.319	41	2352	4.206	ug/l	92
42) 1,2-Dichloroethane	8.252	62	9818	4.913	ug/l	91
43) Isopropyl Acetate	8.289	43	9855	4.597	ug/l #	88
44) Trichloroethene	8.953	130	11434	5.159	ug/l	99
45) 1,2-Dichloropropane	9.227	63	9339	5.042	ug/l	99
46) Dibromomethane	9.313	93	4964	4.946	ug/l	94
47) Bromodichloromethane	9.508	83	13156	5.099	ug/l	95
48) Methyl methacrylate	9.307	41	5166	4.360	ug/l #	85
49) 1,4-Dioxane	9.313	88	909	85.310	ug/l #	66
51) 4-Methyl-2-Pentanone	10.081	43	24629	22.719	ug/l	90
52) Toluene	10.258	92	24776	5.082	ug/l	91
53) t-1,3-Dichloropropene	10.477	75	11969	4.785	ug/l	95
54) cis-1,3-Dichloropropene	9.941	75	14373	4.811	ug/l	91
55) 1,1,2-Trichloroethane	10.654	97	6556	4.739	ug/l	96
56) Ethyl methacrylate	10.520	69	5591	4.282	ug/l #	87
57) 1,3-Dichloropropane	10.800	76	11682	4.954	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.794	63	17683	19.971	ug/l	94
59) 2-Hexanone	10.843	43	19506	24.190	ug/l	91
60) Dibromochloromethane	10.995	129	8241	4.763	ug/l	99
61) 1,2-Dibromoethane	11.105	107	6283	4.767	ug/l	98
64) Tetrachloroethene	10.733	164	10850	5.117	ug/l	96
65) Chlorobenzene	11.526	112	25816	5.117	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.599	131	9279	4.987	ug/l	98
67) Ethyl Benzene	11.605	91	44843	4.946	ug/l	99
68) m/p-Xylenes	11.715	106	33553	9.757	ug/l	97
69) o-Xylene	12.038	106	15193	4.758	ug/l	100
70) Styrene	12.056	104	21442	4.690	ug/l	98
71) Bromoform	12.221	173	4679	4.846	ug/l #	98
73) Isopropylbenzene	12.343	105	42060	4.980	ug/l	100
74) N-amyl acetate	12.154	43	7419	4.374	ug/l #	89
75) 1,1,2,2-Tetrachloroethane	12.593	83	6797	5.016	ug/l	99
76) 1,2,3-Trichloropropane	12.641	75	5056m	4.737	ug/l	
77) Bromobenzene	12.623	156	9561	5.009	ug/l	96
78) n-propylbenzene	12.684	91	50585	5.069	ug/l	98
79) 2-Chlorotoluene	12.769	91	28017	5.094	ug/l	98
80) 1,3,5-Trimethylbenzene	12.824	105	33346	4.897	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.385	75	2030	4.404	ug/l	96
82) 4-Chlorotoluene	12.867	91	28938	5.092	ug/l	100
83) tert-Butylbenzene	13.086	119	29549	4.927	ug/l	97
84) 1,2,4-Trimethylbenzene	13.129	105	32915	4.910	ug/l	99
85) sec-Butylbenzene	13.263	105	43540	5.082	ug/l	100
86) p-Isopropyltoluene	13.379	119	36121	4.896	ug/l	97
87) 1,3-Dichlorobenzene	13.379	146	18752	5.107	ug/l	98
88) 1,4-Dichlorobenzene	13.458	146	18311	5.102	ug/l	92
89) n-Butylbenzene	13.708	91	34034	4.995	ug/l	97
90) Hexachloroethane	13.970	117	7361	5.172	ug/l	86
91) 1,2-Dichlorobenzene	13.751	146	15771	5.012	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.367	75	1239	5.747	ug/l	81
93) 1,2,4-Trichlorobenzene	15.019	180	9028	4.915	ug/l	99
94) Hexachlorobutadiene	15.123	225	5592	5.314	ug/l	99
95) Naphthalene	15.245	128	14668	4.465	ug/l	98
96) 1,2,3-Trichlorobenzene	15.434	180	7426	4.888	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110823\
Data File : VY016253.D
Acq On : 08 Nov 2023 08:59
Operator : SY/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023

Quant Time: Nov 09 00:27:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:22:14 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.801 min Scan# 981
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion:168 Resp: 167284
Ion Ratio Lower Upper
168 100
99 54.2 43.4 65.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#2
Dichlorodifluoromethane
Concen: 5.548 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 85 Resp: 9087
Ion Ratio Lower Upper
85 100
87 32.7 15.6 46.8









#3
Chloromethane
Concen: 5.683 ug/l
RT: 2.119 min Scan# 49
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 50 Resp: 1090
Ion Ratio Lower Upper
50 100
52 31.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#4
Vinyl Chloride
Concen: 5.368 ug/l
RT: 2.253 min Scan# 71
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 62 Resp: 11074
Ion Ratio Lower Upper
62 100
64 27.3 24.3 36.5









#5
Bromomethane
Concen: 5.574 ug/l
RT: 2.661 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 94 Resp: 7550
Ion Ratio Lower Upper
94 100
96 83.4 75.8 113.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#6
Chloroethane
Concen: 5.357 ug/l
RT: 2.802 min Scan# 161
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59


Tgt Ion: 64 Resp: 7051
Ion Ratio Lower Upper
64 100
66 33.6 24.4 36.6





#7
 Trichlorofluoromethane
 Concen: 5.273 ug/l
 RT: 3.131 min Scan# 216
 Delta R.T. -0.007 min
 Lab File: VY016253.D
 Acq: 08 Nov 2023 08:59




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005



Tgt Ion:101 Resp: 15824
 Ion Ratio Lower Upper
 101 100
 103 65.8 51.4 77.0




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
 Supervised By :Mahesh Dadoda 11/09/2023



#8
 Diethyl Ether
 Concen: 4.820 ug/l
 RT: 3.533 min Scan# 281
 Delta R.T. -0.007 min
 Lab File: VY016253.D
 Acq: 08 Nov 2023 08:59

Tgt Ion: 74 Resp: 4323
 Ion Ratio Lower Upper
 74 100
 45 94.1 51.5 154.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.372 ug/l
RT: 3.905 min Scan# 343
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:101 Resp: 9710
Ion Ratio Lower Upper
101 100
85 45.9 36.6 55.0
151 75.6 67.0 100.6




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#10
Methyl Iodide
Concen: 4.808 ug/l
RT: 4.106 min Scan# 375
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion:142 Resp: 9387
Ion Ratio Lower Upper
142 100
127 44.8 37.6 56.4
141 14.3 11.8 17.8









#11
Tert butyl alcohol
Concen: 27.338 ug/l
RT: 4.972 min Scan# 517
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 59 Resp: 327
Ion Ratio Lower Upper
59 100
57 6.7 3.0 4.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#12
1,1-Dichloroethene
Concen: 5.132 ug/l
RT: 3.887 min Scan# 339
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59


Tgt Ion: 96 Resp: 8675
Ion Ratio Lower Upper
96 100
61 138.4 136.4 204.6
98 71.5 49.4 74.2





#13
Acrolein
Concen: 23.018 ug/l
RT: 3.741 min Scan# 315
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 56 Resp: 3324
Ion Ratio Lower Upper
56 100
55 69.3 57.0 85.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#14
Allyl chloride
Concen: 5.046 ug/l
RT: 4.490 min Scan# 438
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59



Tgt Ion: 41 Resp: 12105
Ion Ratio Lower Upper
41 100
39 69.9 58.0 87.0
76 41.3 27.4 41.2#









#15
Acrylonitrile
Concen: 24.433 ug/l
RT: 5.185 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 53 Resp: 869
Ion Ratio Lower Upper
53 100
52 78.3 66.9 100.3
51 35.5 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#16
Acetone
Concen: 29.015 ug/l
RT: 3.954 min Scan# 350
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 43 Resp: 7338
Ion Ratio Lower Upper
43 100
58 28.6 22.7 34.1





#17

Carbon Disulfide

Concen: 5.228 ug/l

RT: 4.210 min Scan# 392

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 2666

Ion Ratio Lower Upper

76 100





78 8.6 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023

Supervised By :Mahesh Dadoda 11/09/2023



#18

Methyl Acetate

Concen: 4.789 ug/l

RT: 4.490 min Scan# 438

Delta R.T. -0.000 min

Lab File: VY016253.D




Acq: 08 Nov 2023 08:59


Tgt Ion: 43 Resp: 6436

Ion Ratio Lower Upper

43 100


74 24.6 17.4 26.0





#19
Methyl tert-butyl Ether
Concen: 4.791 ug/l
RT: 5.234 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 73 Resp: 1960
Ion Ratio Lower Upper
73 100
57 22.0 17.3 25.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#20
Methylene Chloride
Concen: 4.458 ug/l
RT: 4.728 min Scan# 477
Delta R.T. -0.001 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59



Tgt Ion: 84 Resp: 13629
Ion Ratio Lower Upper
84 100
49 109.6 106.9 160.3
51 32.2 33.2 49.8#
86 68.0 51.0 76.6





#21

trans-1,2-Dichloroethene

Concen: 5.285 ug/l

RT: 5.234 min Scan# 50

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 96 Resp: 1028

Ion Ratio Lower Upper

96 100

61 113.9 110.2 165.4





98 58.7 51.8 77.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023

Supervised By :Mahesh Dadoda 11/09/2023



#22

Diisopropyl ether

Concen: 4.880 ug/l

RT: 6.130 min Scan# 707

Delta R.T. -0.007 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion: 45 Resp: 25227




Ion Ratio Lower Upper


45 100

43 55.4 52.2 78.4

87 24.7 21.0 31.6

59 13.6 8.9 13.3#









#23
 Vinyl Acetate
 Concen: 23.200 ug/l
 RT: 6.075 min Scan# 698
 Delta R.T. -0.000 min
 Lab File: VY016253.D
 Acq: 08 Nov 2023 08:59

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Tgt Ion: 43 Resp: 58110
 Ion Ratio Lower Upper
 43 100
 86 10.7 7.5 11.3




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
 Supervised By :Mahesh Dadoda 11/09/2023



#24
 1,1-Dichloroethane
 Concen: 5.218 ug/l
 RT: 6.033 min Scan# 691
 Delta R.T. -0.007 min
 Lab File: VY016253.D
 Acq: 08 Nov 2023 08:59

Tgt Ion: 63 Resp: 16641
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.1 9.4
 100 2.4 1.8 5.5





#25

2-Butanone

Concen: 25.325 ug/l

RT: 6.996 min Scan# 849

Delta R.T. -0.007 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 43 Resp: 11218

Ion Ratio Lower Upper

43 100





72 28.4 20.5 30.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023

Supervised By :Mahesh Dadoda 11/09/2023



#26

2,2-Dichloropropane

Concen: 5.030 ug/l

RT: 6.990 min Scan# 848

Delta R.T. -0.007 min

Lab File: VY016253.D




Acq: 08 Nov 2023 08:59


Tgt Ion: 77 Resp: 14602

Ion Ratio Lower Upper

77 100

97 23.0 10.9 32.9





#27

cis-1,2-Dichloroethene

Concen: 5.124 ug/l

RT: 6.996 min Scan# 849

Delta R.T. -0.007 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 96 Resp: 1074

Ion Ratio Lower Upper

96 100

61 136.2 0.0 301.4





98 65.7 0.0 130.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023

Supervised By :Mahesh Dadoda 11/09/2023



#28

Bromochloromethane

Concen: 5.025 ug/l

RT: 7.350 min Scan# 907

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59




Tgt Ion: 49 Resp: 5931


Ion Ratio Lower Upper

49 100

129 0.4 0.0 4.6

130 83.2 60.7 91.1









#29
Tetrahydrofuran
Concen: 22.566 ug/l
RT: 7.356 min Scan# 909
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 42 Resp: 660
Ion Ratio Lower Upper
42 100
72 50.3 32.8 49.2
71 43.1 31.0 46.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#30
Chloroform
Concen: 5.317 ug/l
RT: 7.520 min Scan# 935
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 83 Resp: 17652
Ion Ratio Lower Upper
83 100
85 69.7 49.7 74.5









#31
Cyclohexane
Concen: 5.628 ug/l
RT: 7.795 min Scan# 981
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 56 Resp: 16940
Ion Ratio Lower Upper
56 100
69 52.5 27.8 41.6
84 100.8 66.4 99.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#32
1,1,1-Trichloroethane
Concen: 5.195 ug/l
RT: 7.715 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 97 Resp: 15809
Ion Ratio Lower Upper
97 100
99 62.4 50.8 76.2
61 41.9 35.1 52.7









#33
1,2-Dichloroethane-d4
Concen: 5.028 ug/l
RT: 8.154 min Scan# 1039
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 65 Resp: 768
Ion Ratio Lower Upper
65 100
67 63.1 0.0 101.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.703 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59


Tgt Ion:114 Resp: 278278
Ion Ratio Lower Upper
114 100
63 19.4 0.0 42.4
88 15.1 0.0 30.6





#35
Dibromofluoromethane
Concen: 5.004 ug/l
RT: 7.728 min Scan# 910
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:113 Resp: 7910
Ion Ratio Lower Upper
113 100
111 101.7 81.3 121.9
192 20.6 16.9 25.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023




#36
1,1-Dichloropropene
Concen: 5.061 ug/l
RT: 7.929 min Scan# 1002
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Tgt Ion: 75 Resp: 13891
Ion Ratio Lower Upper
75 100
110 35.7 17.8 53.5
77 33.7 24.4 36.6





#37
Ethyl Acetate
Concen: 4.776 ug/l
RT: 7.094 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 43 Resp: 529
Ion Ratio Lower Upper
43 100
61 17.4 10.1 15.1
70 10.9 7.1 10.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#38
Carbon Tetrachloride
Concen: 4.991 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59



Tgt Ion:117 Resp: 14219
Ion Ratio Lower Upper
117 100
119 90.9 76.4 114.6
121 27.5 25.8 38.6





#39
Methylcyclohexane
Concen: 4.945 ug/l
RT: 9.197 min Scan# 1210
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion: 83 Resp: 1691
Ion Ratio Lower Upper
83 100
55 67.1 63.0 94.4
98 52.2 37.4 56.2

Manual Integrations




APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#40
Benzene
Concen: 5.056 ug/l
RT: 8.173 min Scan# 1042
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 78 Resp: 38990
Ion Ratio Lower Upper
78 100
77 26.2 19.0 28.6






#41
Methacrylonitrile
Concen: 4.206 ug/l
RT: 7.319 min Scan# 903
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005






Tgt Ion: 41 Resp: 2352
Ion Ratio Lower Upper
41 100
39 64.9 52.1 78.1
67 95.8 66.2 99.2
52 34.1 29.9 44.9


Manual Integrations

APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023




#42
1,2-Dichloroethane
Concen: 4.913 ug/l
RT: 8.252 min Scan# 1055
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59



Tgt Ion: 62 Resp: 9818
Ion Ratio Lower Upper
62 100
98 11.0 0.0 15.8





#43

Isopropyl Acetate

Concen: 4.597 ug/l

RT: 8.289 min Scan# 1060

Delta R.T. 0.006 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion: 43	Resp: 985
Ion Ratio	Lower Upper
43	100
61	33.4 21.1 31.7
87	15.2 9.8 14.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#44

Trichloroethene

Concen: 5.159 ug/l




RT: 8.953 min Scan# 1170


Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion:130	Resp: 11434
Ion Ratio	Lower Upper
130	100
95	99.4 0.0 196.0









#45
1,2-Dichloropropane
Concen: 5.042 ug/l
RT: 9.227 min Scan# 1215
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 63 Resp: 9339
Ion Ratio Lower Upper
63 100
65 30.2 23.8 35.6




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#46
Dibromomethane
Concen: 4.946 ug/l
RT: 9.313 min Scan# 1229
Delta R.T. -0.007 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 93 Resp: 4964
Ion Ratio Lower Upper
93 100
95 81.7 66.6 100.0
174 96.9 84.8 127.2









#47
Bromodichloromethane
Concen: 5.099 ug/l
RT: 9.508 min Scan# 1261
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 13150
Ion Ratio Lower Upper
83 100
85 58.3 50.1 75.1
127 9.4 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#48
Methyl methacrylate
Concen: 4.360 ug/l
RT: 9.307 min Scan# 1228
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 41 Resp: 5166
Ion Ratio Lower Upper
41 100
69 99.1 64.5 96.7#
39 57.1 50.8 76.2





#49

1,4-Dioxane

Concen: 85.310 ug/l

RT: 9.313 min Scan# 1229

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion: 88 Resp: 909
 Ion Ratio Lower Upper
 88 100
 43 40.3 0.0 0.0
 58 59.2 30.9 46.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
 Supervised By :Mahesh Dadoda 11/09/2023



#50

Toluene-d8

Concen: 4.823 ug/l




RT: 10.191 min Scan# 1373


Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion: 98 Resp: 30918
 Ion Ratio Lower Upper
 98 100
 100 66.3 50.1 75.1









#51
4-Methyl-2-Pentanone
Concen: 22.719 ug/l
RT: 10.081 min Scan# 1355
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 24629
Ion Ratio Lower Upper
43 100
58 42.0 29.0 43.6




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#52
Toluene
Concen: 5.082 ug/l
RT: 10.258 min Scan# 1384
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59


Tgt Ion: 92 Resp: 24776
Ion Ratio Lower Upper
92 100
91 166.9 143.8 215.6





#53
t-1,3-Dichloropropene
Concen: 4.785 ug/l
RT: 10.477 min Scan# 1420
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument : MSVOA_Y
ClientSampleId : VSTDIC005



Tgt Ion: 75 Resp: 11969
Ion Ratio Lower Upper
75 100
77 28.4 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#54
cis-1,3-Dichloropropene
Concen: 4.811 ug/l
RT: 9.941 min Scan# 1332
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 75 Resp: 14373
Ion Ratio Lower Upper
75 100
77 31.3 24.2 36.4
39 43.5 42.2 63.4









#55
1,1,2-Trichloroethane
Concen: 4.739 ug/l
RT: 10.654 min Scan# 1449
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument : MSVOA_Y
ClientSampleId : VSTDIC005

Tgt Ion	Ratio	Lower	Upper
97	100		
83	85.5	72.0	108.0
85	55.4	45.4	68.0
99	58.2	50.3	75.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#56
Ethyl methacrylate
Concen: 4.282 ug/l
RT: 10.520 min Scan# 1427
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion	Ratio	Lower	Upper
69	100		
41	63.3	57.6	86.4
39	32.3	34.1	51.1









#57
1,3-Dichloropropane
Concen: 4.954 ug/l
RT: 10.800 min Scan# 1473
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument : MSVOA_Y
ClientSampleId : VSTDIC005

Tgt Ion: 76 Resp: 1168
Ion Ratio Lower Upper
76 100
78 31.9 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#58
2-Chloroethyl Vinyl ether
Concen: 19.971 ug/l
RT: 9.794 min Scan# 1308
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 63 Resp: 17683
Ion Ratio Lower Upper
63 100
106 29.8 21.2 31.8









#59
2-Hexanone
Concen: 24.190 ug/l
RT: 10.843 min Scan# 1480
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 19500
Ion Ratio Lower Upper
43 100
58 56.3 25.1 75.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#60
Dibromochloromethane
Concen: 4.763 ug/l
RT: 10.995 min Scan# 1505
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59


Tgt Ion:129 Resp: 8241
Ion Ratio Lower Upper
129 100
127 78.2 38.5 115.5





#61
1,2-Dibromoethane
Concen: 4.767 ug/l
RT: 11.105 min Scan# 11105
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion:107 Resp: 628
Ion Ratio Lower Upper
107 100
109 93.0 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#62
4-Bromofluorobenzene
Concen: 4.877 ug/l
RT: 12.495 min Scan# 1751
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59



Tgt Ion: 95 Resp: 10155
Ion Ratio Lower Upper
95 100
174 87.3 0.0 178.2
176 87.4 0.0 173.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.501 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:117	Resp: 23612
Ion Ratio	Lower Upper
117 100	
82 52.9	43.4 65.0
119 32.7	26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#64

Tetrachloroethene

Concen: 5.117 ug/l




RT: 10.733 min Scan# 1462


Delta R.T. 0.006 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion:164	Resp: 10850
Ion Ratio	Lower Upper
164 100	
166 129.9	103.0 154.6
129 93.3	77.2 115.8
131 95.6	69.6 104.4









#65
Chlorobenzene
Concen: 5.117 ug/l
RT: 11.526 min Scan# 11592
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:112 Resp: 25810
Ion Ratio Lower Upper
112 100
114 30.5 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 4.987 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion:131 Resp: 9279
Ion Ratio Lower Upper
131 100
133 96.2 47.5 142.6
119 64.9 31.6 94.7





#67

Ethyl Benzene

Concen: 4.946 ug/l

RT: 11.605 min Scan# 1605

Delta R.T. 0.006 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 4484

Ion Ratio Lower Upper

91 100





106 31.5 24.6 36.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023

Supervised By :Mahesh Dadoda 11/09/2023



#68

m/p-Xylenes

Concen: 9.757 ug/l

RT: 11.715 min Scan# 1623

Delta R.T. -0.000 min

Lab File: VY016253.D




Acq: 08 Nov 2023 08:59


Tgt Ion:106 Resp: 33553

Ion Ratio Lower Upper

106 100

91 202.7 166.1 249.1









#69
o-Xylene
Concen: 4.758 ug/l
RT: 12.038 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:106 Resp: 1519
Ion Ratio Lower Upper
106 100
91 216.7 108.1 324.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#70
Styrene
Concen: 4.690 ug/l
RT: 12.056 min Scan# 1679
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59


Tgt Ion:104 Resp: 21442
Ion Ratio Lower Upper
104 100
78 49.6 41.5 62.3
103 54.6 44.6 67.0





#71
Bromoform
Concen: 4.846 ug/l
RT: 12.221 min Scan# 1705
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:173 Resp: 4679
Ion Ratio Lower Upper
173 100
175 47.7 24.6 73.8
254 0.0 0.2 0.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.434 min Scan# 1905
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59



Tgt Ion:152 Resp: 104288
Ion Ratio Lower Upper
152 100
115 58.5 28.8 86.5
150 158.8 0.0 348.8









#73
Isopropylbenzene
Concen: 4.980 ug/l
RT: 12.343 min Scan# 1726
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:105 Resp: 42060
Ion Ratio Lower Upper
105 100
120 26.4 13.3 39.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#74
N-ethyl acetate
Concen: 4.374 ug/l
RT: 12.154 min Scan# 1695
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59


Tgt Ion: 43 Resp: 7419
Ion Ratio Lower Upper
43 100
70 47.1 29.7 44.5#
55 27.5 18.3 27.5#
61 22.5 17.5 26.3





#75
1,1,2,2-Tetrachloroethane
Concen: 5.016 ug/l
RT: 12.593 min Scan# 1767
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59




Instrument : MSVOA_Y
ClientSampleId : VSTDIC005




Tgt Ion: 83 Resp: 679
Ion Ratio Lower Upper
83 100
131 10.1 5.4 16.2
85 62.8 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023




#76
1,2,3-Trichloropropane
Concen: 4.737 ug/l m
RT: 12.641 min Scan# 1775
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59



Tgt Ion: 75 Resp: 5056
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 5.009 ug/l

RT: 12.623 min Scan# 1772

Delta R.T. 0.006 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	156	Resp:	956
Ion Ratio	Lower	Upper	
156	100		
77	178.4	92.8	278.6
158	98.0	48.4	145.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#78

n-propylbenzene

Concen: 5.069 ug/l




RT: 12.684 min Scan# 1782


Delta R.T. 0.006 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion:	91	Resp:	50585
Ion Ratio	Lower	Upper	
91	100		
120	23.6	11.5	34.4









#79
2-Chlorotoluene
Concen: 5.094 ug/l
RT: 12.769 min Scan# 1795
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 2801
Ion Ratio Lower Upper
91 100
126 35.7 17.3 51.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#80
1,3,5-Trimethylbenzene
Concen: 4.897 ug/l
RT: 12.824 min Scan# 1805
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion:105 Resp: 33346
Ion Ratio Lower Upper
105 100
120 49.1 23.9 71.7









#81
trans-1,4-Dichloro-2-butene
Concen: 4.404 ug/l
RT: 12.385 min Scan# 1733
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 75 Resp: 2030
Ion Ratio Lower Upper
75 100
53 93.2 77.8 116.8
89 43.8 37.4 56.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#82
4-Chlorotoluene
Concen: 5.092 ug/l
RT: 12.867 min Scan# 1812
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 91 Resp: 28938
Ion Ratio Lower Upper
91 100
126 34.4 17.2 51.6





#83

tert-Butylbenzene

Concen: 4.927 ug/l

RT: 13.086 min Scan# 1847

Delta R.T. 0.006 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	119	Resp:	29549
Ion Ratio	Lower	Upper	
119	100		
91	63.6	32.9	98.6
134	27.8	13.0	38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#84

1,2,4-Trimethylbenzene

Concen: 4.910 ug/l




RT: 13.129 min Scan# 1855


Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion:	105	Resp:	32915
Ion Ratio	Lower	Upper	
105	100		
120	44.8	22.2	66.6





#85

sec-Butylbenzene

Concen: 5.082 ug/l

RT: 13.263 min Scan# 1877

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion:105 Resp: 43540

Ion Ratio Lower Upper

105 100





134 20.3 10.2 30.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023

Supervised By :Mahesh Dadoda 11/09/2023



#86

p-Isopropyltoluene

Concen: 4.896 ug/l

RT: 13.379 min Scan# 1896

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59




Tgt Ion:119 Resp: 36121


Ion Ratio Lower Upper

119 100

134 27.6 13.3 39.8

91 22.6 12.4 37.4









#87
1,3-Dichlorobenzene
Concen: 5.107 ug/l
RT: 13.379 min Scan# 1895
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:146 Resp: 1875
Ion Ratio Lower Upper
146 100
111 39.8 19.4 58.2
148 65.0 31.9 95.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#88
1,4-Dichlorobenzene
Concen: 5.102 ug/l
RT: 13.458 min Scan# 1909
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion:146 Resp: 18311
Ion Ratio Lower Upper
146 100
111 45.8 18.9 56.7
148 67.7 32.1 96.5









#89
n-Butylbenzene
Concen: 4.995 ug/l
RT: 13.708 min Scan# 1949
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 34034
Ion Ratio Lower Upper
91 100
92 53.1 25.4 76.0
134 26.6 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#90
Hexachloroethane
Concen: 5.172 ug/l
RT: 13.970 min Scan# 1993
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion:117 Resp: 7361
Ion Ratio Lower Upper
117 100
201 70.1 41.6 124.8









#91
1,2-Dichlorobenzene
Concen: 5.012 ug/l
RT: 13.751 min Scan# 1956
Delta R.T. 0.006 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:146 Resp: 1577
Ion Ratio Lower Upper
146 100
111 41.0 19.7 59.0
148 62.8 32.0 96.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 5.747 ug/l
RT: 14.367 min Scan# 2058
Delta R.T. -0.000 min
Lab File: VY016253.D
Acq: 08 Nov 2023 08:59

Tgt Ion: 75 Resp: 1239
Ion Ratio Lower Upper
75 100
155 61.3 41.5 124.6
157 88.0 51.8 155.5





#93

1,2,4-Trichlorobenzene

Concen: 4.915 ug/l

RT: 15.019 min Scan# 2165

Delta R.T. 0.006 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:	180	Resp:	902
Ion Ratio	Lower	Upper	
180	100		
182	93.3	47.0	141.2
145	30.3	14.5	43.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#94

Hexachlorobutadiene

Concen: 5.314 ug/l




RT: 15.123 min Scan# 2182


Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion:	225	Resp:	5592
Ion Ratio	Lower	Upper	
225	100		
223	62.5	31.9	95.9
227	65.4	33.0	98.9





#95

Naphthalene

Concen: 4.465 ug/l

RT: 15.245 min Scan# 2202

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:128	Resp:	1466
Ion Ratio	Lower	Upper
128	100	
127	12.8	10.8 16.2
129	11.8	8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 11/09/2023
Supervised By :Mahesh Dadoda 11/09/2023



#96

1,2,3-Trichlorobenzene

Concen: 4.888 ug/l

RT: 15.434 min Scan# 2233

Delta R.T. -0.000 min

Lab File: VY016253.D

Acq: 08 Nov 2023 08:59

Tgt Ion:180	Resp:	7426
Ion Ratio	Lower	Upper
180	100	
182	90.0	47.3 141.8
145	33.1	15.4 46.2

