

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
 Data File : VY020227.D
 Acq On : 08 Nov 2024 12:51
 Operator : SY/MD
 Sample : P4739-03
 Misc : 4.78g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :


Quant Time: Nov 08 22:17:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	189977	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	398216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	394058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	200458	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	142388	60.053	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 120.100%	
35) Dibromofluoromethane	7.634	113	133839	52.862	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.720%	
50) Toluene-d8	10.103	98	514692	51.071	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.140%	
62) 4-Bromofluorobenzene	12.402	95	268929	82.155	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 164.320%#	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.848	59	1430	8.385	ug/l	99
13) Acrolein	3.653	56	296	1.394	ug/l	87
16) Acetone	3.873	43	23753	43.078	ug/l	# 86
20) Methylene Chloride	4.598	84	7071	2.855	ug/l	# 87
25) 2-Butanone	6.903	43	6251	8.363	ug/l	90
31) Cyclohexane	7.707	56	6968	1.619	ug/l	# 45
36) 1,1-Dichloropropene	7.707	75	19790	5.060	ug/l	# 48
38) Carbon Tetrachloride	7.713	117	22349	5.682	ug/l	# 17
39) Methylcyclohexane	9.079	83	11334	2.283	ug/l	# 12
48) Methyl methacrylate	9.250	41	1828	1.101	ug/l	# 7
49) 1,4-Dioxane	9.225	88	23	1.510	ug/l	# 1
51) 4-Methyl-2-Pentanone	10.000	43	1927	1.043	ug/l	# 66
55) 1,1,2-Trichloroethane	10.542	97	74144	40.294	ug/l	# 22
56) Ethyl methacrylate	10.445	69	29048	11.005	ug/l	# 1
59) 2-Hexanone	10.731	43	63366	47.838	ug/l	# 25
74) N-amyl acetate	12.048	43	233446	58.318	ug/l	# 37
75) 1,1,2,2-Tetrachloroethane	12.481	83	89272	31.876	ug/l	# 92
76) 1,2,3-Trichloropropane	12.402	75	104572	55.651	ug/l	# 100
78) n-propylbenzene	12.566	91	26035	1.306	ug/l	# 53
79) 2-Chlorotoluene	12.664	91	21815	1.972	ug/l	# 54
80) 1,3,5-Trimethylbenzene	12.895	105	14039	1.092	ug/l	71
81) trans-1,4-Dichloro-2-b...	12.402	75	104572	119.702	ug/l	# 35
82) 4-Chlorotoluene	12.731	91	19349	1.715	ug/l	# 34
83) tert-Butylbenzene	12.999	119	159327	14.173	ug/l	81
84) 1,2,4-Trimethylbenzene	12.895	105	14039	1.103	ug/l	76
85) sec-Butylbenzene	13.176	105	153142	8.874	ug/l	# 20
86) p-Isopropyltoluene	13.444	119	100358	7.247	ug/l	# 74
89) n-Butylbenzene	13.615	91	58899	4.243	ug/l	86
90) Hexachloroethane	13.895	117	62419	22.790	ug/l	# 12
92) 1,2-Dibromo-3-Chloropr...	14.298	75	3386	7.972	ug/l	# 1
93) 1,2,4-Trichlorobenzene	14.919	180	6276	2.043	ug/l	# 1
95) Naphthalene	15.115	128	41153	7.166	ug/l	# 1
96) 1,2,3-Trichlorobenzene	15.297	180	24955	9.569	ug/l	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed





Instrument :
MSVOA_Y
ClientSampleId :



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:168 Resp: 189977
Ion Ratio Lower Upper
168 100
99 67.6 39.1 58.7#



#11
Tert butyl alcohol
Concen: 8.385 ug/l
RT: 4.848 min Scan# 513
Delta R.T. -0.012 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51

Tgt Ion: 59 Resp: 1430
Ion Ratio Lower Upper
59 100
57 4.3 3.0 4.6





#13

Acrolein

Concen: 1.394 ug/l

RT: 3.653 min Scan# 317

Delta R.T. 0.000 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 56 Resp: 296

Ion Ratio Lower Upper

56 100

55 59.1 55.7 83.5



#16

Acetone

Concen: 43.078 ug/l

RT: 3.873 min Scan# 353

Delta R.T. 0.000 min

Lab File: VY020227.D




Acq: 08 Nov 2024 12:51


Tgt Ion: 43 Resp: 23753

Ion Ratio Lower Upper

43 100


58 31.1 31.7 47.5#








#20
Methylene Chloride
Concen: 2.855 ug/l
RT: 4.598 min Scan# 41
Delta R.T. -0.018 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:	84	Resp:	7071
Ion	Ratio	Lower	Upper
84	100		
49	101.9	84.4	126.6
51	25.1	26.0	39.0#
86	39.9	49.5	74.3#



#25
2-Butanone
Concen: 8.363 ug/l
RT: 6.903 min Scan# 850
Delta R.T. 0.013 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51

Tgt Ion:	43	Resp:	6251
Ion	Ratio	Lower	Upper
43	100		
72	28.2	27.3	40.9









#31
Cyclohexane
Concen: 1.619 ug/l
RT: 7.707 min Scan# 981
Delta R.T. 0.006 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 56 Resp: 6968
Ion Ratio Lower Upper
56 100
69 130.7 28.7 43.1#
84 112.3 77.7 116.5



#33
1,2-Dichloroethane-d4
Concen: 60.053 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51

Tgt Ion: 65 Resp: 142388
Ion Ratio Lower Upper
65 100
67 51.9 0.0 109.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:114 Resp: 398216

Ion Ratio Lower Upper

114 100

63 22.8 0.0 35.0

88 15.5 0.0 27.2



#35

Dibromofluoromethane

Concen: 52.862 ug/l

RT: 7.634 min Scan# 970

Delta R.T. 0.000 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion:113 Resp: 133839


Ion Ratio Lower Upper

113 100

111 103.3 82.2 123.4

192 16.4 15.9 23.9





#36

1,1-Dichloropropene

Concen: 5.060 ug/l

RT: 7.707 min Scan# 982

Delta R.T. -0.128 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 75 Resp: 19790

Ion Ratio Lower Upper

75 100

110 9.1 19.0 57.0#

77 0.0 25.3 37.9#



#38

Carbon Tetrachloride

Concen: 5.682 ug/l

RT: 7.713 min Scan# 983

Delta R.T. -0.103 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion: 117 Resp: 22349


Ion Ratio Lower Upper

117 100

119 5.5 75.8 113.6#

121 0.0 24.0 36.0#





#39

Methylcyclohexane

Concen: 2.283 ug/l

RT: 9.079 min Scan# 11

Delta R.T. -0.030 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 83 Resp: 11334

Ion Ratio Lower Upper

83 100

55 152.5 51.5 77.3#

98 11.4 40.5 60.7#



#48

Methyl methacrylate

Concen: 1.101 ug/l

RT: 9.250 min Scan# 1235

Delta R.T. 0.031 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion: 41 Resp: 1828


Ion Ratio Lower Upper

41 100

69 0.0 84.1 126.1#

39 0.0 43.6 65.4#





#49

1,4-Dioxane

Concen: 1.510 ug/l

RT: 9.225 min Scan# 1

Delta R.T. -0.006 min


Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y

ClientSampleId :






Tgt Ion: 88 Resp: 23

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 473.9 49.3 73.9#



#50

Toluene-d8

Concen: 51.071 ug/l

RT: 10.103 min Scan# 1375

Delta R.T. -0.006 min

Lab File: VY020227.D




Acq: 08 Nov 2024 12:51


Tgt Ion: 98 Resp: 514692

Ion Ratio Lower Upper

98 100

100 63.4 52.0 78.0





#51

4-Methyl-2-Pentanone

Concen: 1.043 ug/l

RT: 10.000 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 43 Resp: 1927

Ion Ratio Lower Upper

43 100

58 22.2 35.7 53.5#



#55

1,1,2-Trichloroethane

Concen: 40.294 ug/l

RT: 10.542 min Scan# 1447

Delta R.T. -0.030 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Tgt Ion: 97 Resp: 74144




Ion Ratio Lower Upper


97 100

83 8.2 67.5 101.3#

85 3.3 42.9 64.3#

99 0.4 46.6 70.0#









#56
Ethyl methacrylate
Concen: 11.005 ug/l
RT: 10.445 min Scan# 1430
Delta R.T. 0.006 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 69 Resp: 29048
Ion Ratio Lower Upper
69 100
41 204.5 42.3 63.5#
39 104.6 22.9 34.3#



#59
2-Hexanone
Concen: 47.838 ug/l
RT: 10.731 min Scan# 1478
Delta R.T. -0.030 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51

Tgt Ion: 43 Resp: 63366
Ion Ratio Lower Upper
43 100
58 4.4 31.4 94.0#





#62

4-Bromofluorobenzene

Concen: 82.155 ug/l

RT: 12.402 min Scan# 1

Delta R.T. -0.006 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 95 Resp: 268929

Ion Ratio Lower Upper

95 100

174 49.4 0.0 175.6

176 47.0 0.0 171.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 1590

Delta R.T. -0.006 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion: 117 Resp: 394058


Ion Ratio Lower Upper

117 100

82 55.6 42.4 63.6


119 31.3 25.9 38.9








#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.347 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion:152 Resp: 200458
Ion Ratio Lower Upper
152 100
115 50.0 28.2 84.7
150 123.7 0.0 345.6




#74
N-ethyl acetate
Concen: 58.318 ug/l
RT: 12.048 min Scan# 1694
Delta R.T. -0.024 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51



Tgt Ion: 43 Resp: 233446
Ion Ratio Lower Upper
43 100
70 0.0 41.4 62.2#
55 0.0 23.3 34.9#
61 0.0 22.6 33.8#





#75

1,1,2,2-Tetrachloroethane

Concen: 31.876 ug/l

RT: 12.481 min Scan# 1

Delta R.T. -0.030 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 83 Resp: 89272

Ion Ratio Lower Upper

83 100

131 0.0 5.0 14.9#

85 68.2 32.0 96.0



#76

1,2,3-Trichloropropane

Concen: 55.651 ug/l

RT: 12.402 min Scan# 1752

Delta R.T. -0.158 min

Lab File: VY020227.D




Acq: 08 Nov 2024 12:51


Tgt Ion: 75 Resp: 104572

Ion Ratio Lower Upper

75 100

77 12.1 0.0 0.0#









#78
n-propylbenzene
Concen: 1.306 ug/l
RT: 12.566 min Scan# 1779
Delta R.T. -0.030 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 26035
Ion Ratio Lower Upper
91 100
120 0.0 11.5 34.4#



#79
2-Chlorotoluene
Concen: 1.972 ug/l
RT: 12.664 min Scan# 1795
Delta R.T. -0.018 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51

Tgt Ion: 91 Resp: 21815
Ion Ratio Lower Upper
91 100
126 8.7 17.8 53.5#





#80

1,3,5-Trimethylbenzene

Concen: 1.092 ug/l

RT: 12.895 min Scan# 11

Delta R.T. 0.159 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:105 Resp: 14039

Ion Ratio Lower Upper

105 100

120 30.3 25.1 75.2



#81

trans-1,4-Dichloro-2-butene

Concen: 119.702 ug/l

RT: 12.402 min Scan# 1752

Delta R.T. 0.098 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion: 75 Resp: 104572


Ion Ratio Lower Upper

75 100

53 25.4 66.8 100.2#

89 1.2 35.8 53.6#





#82

4-Chlorotoluene

Concen: 1.715 ug/l

RT: 12.731 min Scan# 1806

Delta R.T. -0.048 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 91 Resp: 19349

Ion Ratio Lower Upper

91 100

126 73.8 17.6 52.9#



#83

tert-Butylbenzene

Concen: 14.173 ug/l

RT: 12.999 min Scan# 1850

Delta R.T. 0.000 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion: 119 Resp: 159327


Ion Ratio Lower Upper

119 100

91 81.8 31.3 93.8

134 24.5 13.4 40.1









#84
1,2,4-Trimethylbenzene
Concen: 1.103 ug/l
RT: 12.895 min Scan# 1833
Delta R.T. -0.152 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:105 Resp: 14039
Ion Ratio Lower Upper
105 100
120 30.3 23.2 69.6



#85
sec-Butylbenzene
Concen: 8.874 ug/l
RT: 13.176 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VY020227.D
Acq: 08 Nov 2024 12:51

Tgt Ion:105 Resp: 153142
Ion Ratio Lower Upper
105 100
134 57.0 10.1 30.1#





#86

p-Isopropyltoluene

Concen: 7.247 ug/l

RT: 13.444 min Scan# 1923

Delta R.T. 0.153 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion:119 Resp: 100358

Ion Ratio Lower Upper

119 100

134 27.3 13.1 39.3

91 47.5 10.7 32.1#



#89

n-Butylbenzene

Concen: 4.243 ug/l

RT: 13.615 min Scan# 1951

Delta R.T. -0.006 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion: 91 Resp: 58899


Ion Ratio Lower Upper

91 100

92 44.8 26.6 79.7

134 34.9 12.9 38.7





#90

Hexachloroethane

Concen: 22.790 ug/l

RT: 13.895 min Scan# 1994

Delta R.T. 0.013 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:117 Resp: 62419

Ion Ratio Lower Upper

117 100

201 0.0 37.4 112.2#



#92

1,2-Dibromo-3-Chloropropane

Concen: 7.972 ug/l

RT: 14.298 min Scan# 2063

Delta R.T. 0.025 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Tgt Ion: 75 Resp: 3386


Ion Ratio Lower Upper

75 100

155 0.0 48.6 145.8#

157 1.4 62.7 188.2#





#93

1,2,4-Trichlorobenzene

Concen: 2.043 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. 0.000 min


Lab File: VY020227.D

Acq: 08 Nov 2024 12:51




Instrument :

MSVOA_Y

ClientSampleId :



Tgt Ion:180	Resp:	6276
Ion Ratio	Lower	Upper
180	100	
182	408.6	47.9 143.8#
145	5776.2	14.2 42.6#



#95

Naphthalene

Concen: 7.166 ug/l




RT: 15.115 min Scan# 2197


Delta R.T. -0.030 min

Lab File: VY020227.D

Acq: 08 Nov 2024 12:51


Tgt Ion:128	Resp:	41153
Ion Ratio	Lower	Upper
128	100	
127	118.7	10.6 16.0#
129	60.7	8.6 13.0#





#96
 1,2,3-Trichlorobenzene
 Concen: 9.569 ug/l
 RT: 15.297 min Scan# 21
 Delta R.T. -0.030 min
 Lab File: VY020227.D
 Acq: 08 Nov 2024 12:51

Instrument :
 MSVOA_Y
 ClientSampleId :



Tgt Ion:180 Resp: 24955
 Ion Ratio Lower Upper
 180 100
 182 18.9 47.6 142.9#
 145 769.9 15.4 46.1#

