

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
 Data File : VY020226.D
 Acq On : 08 Nov 2024 12:28
 Operator : SY/MD
 Sample : P4720-02
 Misc : 4.93g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JC-701-VOC-01

Quant Time: Nov 08 13:22:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	24428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	44287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	37922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	10758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	17525	57.482	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.960%
35) Dibromofluoromethane	7.634	113	15726	55.850	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.700%
50) Toluene-d8	10.109	98	54403	48.539	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.080%
62) 4-Bromofluorobenzene	12.407	95	14440	39.665	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	79.320%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
Data File : VY020226.D
Acq On : 08 Nov 2024 12:28
Operator : SY/MD
Sample : P4720-02
Misc : 4.93g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JC-701-VOC-01

Quant Time: Nov 08 13:22:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

