

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
 Data File : VY020232.D
 Acq On : 08 Nov 2024 14:48
 Operator : SY/MD
 Sample : P4771-05
 Misc : 7.55g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P-1

Quant Time: Nov 08 22:19:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	197681	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	425026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	384906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	138261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	144643	58.627	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.260%
35) Dibromofluoromethane	7.634	113	137292	50.806	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.620%
50) Toluene-d8	10.109	98	509992	47.413	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.820%
62) 4-Bromofluorobenzene	12.408	95	164567	47.102	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.200%
Target Compounds						
20) Methylene Chloride	4.622	84	12878	4.996	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
Data File : VY020232.D
Acq On : 08 Nov 2024 14:48
Operator : SY/MD
Sample : P4771-05
Misc : 7.55g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P-1

Quant Time: Nov 08 22:19:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

