

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
 Data File : VY020234.D
 Acq On : 08 Nov 2024 15:35
 Operator : SY/MD
 Sample : P4718-02
 Misc : 7.96g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WB-307-SB02

Quant Time: Nov 08 22:20:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	200262	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	428332	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	399084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	136217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	157720	63.103	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.200%
35) Dibromofluoromethane	7.634	113	140994	51.773	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.540%
50) Toluene-d8	10.103	98	531619	49.042	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.408	95	167289	47.512	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	95.020%
Target Compounds						
16) Acetone	3.885	43	4260	7.329	ug/l	# 81
20) Methylene Chloride	4.616	84	12182	4.665	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
Data File : VY020234.D
Acq On : 08 Nov 2024 15:35
Operator : SY/MD
Sample : P4718-02
Misc : 7.96g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WB-307-SB02

Quant Time: Nov 08 22:20:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

