

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
 Data File : VY020236.D
 Acq On : 08 Nov 2024 17:49
 Operator : SY/MD
 Sample : P4737-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP2

Quant Time: Nov 08 22:20:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	200069	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	420606	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	385502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	140838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	143119	57.317	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.640%
35) Dibromofluoromethane	7.640	113	138146	51.659	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.320%
50) Toluene-d8	10.109	98	532013	49.980	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.960%
62) 4-Bromofluorobenzene	12.407	95	164118	47.467	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	94.940%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110824\
Data File : VY020236.D
Acq On : 08 Nov 2024 17:49
Operator : SY/MD
Sample : P4737-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP2

Quant Time: Nov 08 22:20:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

