

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110923\
 Data File : VY016273.D
 Acq On : 09 Nov 2023 11:53
 Operator : SY/MD
 Sample : 05317-03
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-21148

Quant Time: Nov 09 22:56:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	106421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	185273	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	174641	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	80745	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	62693	64.456	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.920%
35) Dibromofluoromethane	7.728	113	56987	54.146	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.300%
50) Toluene-d8	10.191	98	215885	50.583	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.160%
62) 4-Bromofluorobenzene	12.489	95	74862	54.006	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110923\
Data File : VY016273.D
Acq On : 09 Nov 2023 11:53
Operator : SY/MD
Sample : 05317-03
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-21148

Quant Time: Nov 09 22:56:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

