

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
 Data File : VY016301.D
 Acq On : 10 Nov 2023 12:24
 Operator : SY/MD
 Sample : 05345-01
 Misc : 4.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NEW-PARKING-LOT-STOCKPILE-A

Quant Time: Nov 10 22:19:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	100344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	180237	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	175385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	82186	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	64402	70.223	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.440%	
35) Dibromofluoromethane	7.728	113	55835	54.534	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.060%	
50) Toluene-d8	10.191	98	211598	50.963	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.920%	
62) 4-Bromofluorobenzene	12.489	95	76888	57.018	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 114.040%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
Data File : VY016301.D
Acq On : 10 Nov 2023 12:24
Operator : SY/MD
Sample : 05345-01
Misc : 4.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NEW-PARKING-LOT-STOCKPILE-A

Quant Time: Nov 10 22:19:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

