

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
 Data File : VY016305.D
 Acq On : 10 Nov 2023 13:58
 Operator : SY/MD
 Sample : 05351-01
 Misc : 4.12 g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-241

Quant Time: Nov 10 22:21:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	100699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	178056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	170093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	78625	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	61042	66.325	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.640%
35) Dibromofluoromethane	7.734	113	55183	54.557	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.120%
50) Toluene-d8	10.191	98	208274	50.777	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.560%
62) 4-Bromofluorobenzene	12.489	95	72257	54.240	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
Data File : VY016305.D
Acq On : 10 Nov 2023 13:58
Operator : SY/MD
Sample : 05351-01
Misc : 4.12 g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-241

Quant Time: Nov 10 22:21:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

