

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
 Data File : VY016309.D
 Acq On : 10 Nov 2023 17:13
 Operator : SY/MD
 Sample : 05379-07
 Misc : 4.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-4-VOC

Quant Time: Nov 10 22:22:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	44127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	76279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	72187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	31745	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	27322	67.745	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.500%
35) Dibromofluoromethane	7.728	113	24710	57.026	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.060%
50) Toluene-d8	10.191	98	89238	50.785	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.560%
62) 4-Bromofluorobenzene	12.489	95	29609	51.882	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
Data File : VY016309.D
Acq On : 10 Nov 2023 17:13
Operator : SY/MD
Sample : 05379-07
Misc : 4.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-4-VOC

Quant Time: Nov 10 22:22:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

