

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
 Data File : VY016311.D
 Acq On : 10 Nov 2023 18:00
 Operator : SY/MD
 Sample : 05380-03
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-7-VOC

Quant Time: Nov 10 22:23:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	96749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	175498	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	167783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	79413	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	62980	71.224	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.440%
35) Dibromofluoromethane	7.734	113	54902	55.071	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.140%
50) Toluene-d8	10.191	98	206225	51.010	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.020%
62) 4-Bromofluorobenzene	12.495	95	72935	55.547	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	111.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111023\
Data File : VY016311.D
Acq On : 10 Nov 2023 18:00
Operator : SY/MD
Sample : 05380-03
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-7-VOC

Quant Time: Nov 10 22:23:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

