

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111323\
 Data File : VY016326.D
 Acq On : 13 Nov 2023 16:32
 Operator : SY/MD
 Sample : 05377-03
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 289

Quant Time: Nov 14 04:38:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	101494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	183399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	175812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	81385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	64594	69.634	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.260%	
35) Dibromofluoromethane	7.734	113	57594	55.282	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.560%	
50) Toluene-d8	10.191	98	213053	50.429	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.860%	
62) 4-Bromofluorobenzene	12.495	95	75037	54.686	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 109.380%	
Target Compounds						
16) Acetone	3.966	43	1851	12.063	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111323\
Data File : VY016326.D
Acq On : 13 Nov 2023 16:32
Operator : SY/MD
Sample : 05377-03
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
289

Quant Time: Nov 14 04:38:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

