

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111323\
 Data File : VY016328.D
 Acq On : 13 Nov 2023 17:19
 Operator : SY/MD
 Sample : 05366-01
 Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P-13(10-15)

Quant Time: Nov 14 04:38:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	97549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	171232	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	165969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	80092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	60548	67.912	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.820%
35) Dibromofluoromethane	7.734	113	53915	55.428	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.860%
50) Toluene-d8	10.197	98	203324	51.546	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.100%
62) 4-Bromofluorobenzene	12.495	95	73355	57.258	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	114.520%
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	12367	83.858	ug/l #	86
17) Carbon Disulfide	4.216	76	8683	2.919	ug/l	100
20) Methylene Chloride	4.735	84	12041	8.681	ug/l	89
40) Benzene	8.173	78	11575	2.439	ug/l	97
95) Naphthalene	15.245	128	35349	14.012	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111323\
Data File : VY016328.D
Acq On : 13 Nov 2023 17:19
Operator : SY/MD
Sample : 05366-01
Misc : 5.71g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P-13(10-15)

Quant Time: Nov 14 04:38:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

