

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111323\
 Data File : VY016331.D
 Acq On : 13 Nov 2023 18:29
 Operator : SY/MD
 Sample : BLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLK

Quant Time: Nov 14 04:39:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	102323	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	180804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	178176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	83631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	62409	66.734	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	133.460%	
35) Dibromofluoromethane	7.734	113	56604	55.112	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	110.220%	
50) Toluene-d8	10.197	98	215015	51.624	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	103.240%	
62) 4-Bromofluorobenzene	12.495	95	79362	58.668	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	117.340%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.741	84	12203	8.261	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111323\
Data File : VY016331.D
Acq On : 13 Nov 2023 18:29
Operator : SY/MD
Sample : BLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLK

Quant Time: Nov 14 04:39:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

