

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111325\
 Data File : VY023779.D
 Acq On : 13 Nov 2025 12:14
 Operator : SY/MD
 Sample : Q3615-01
 Misc : 5.50g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 East Bathhouse- Concrete

Quant Time: Nov 13 13:32:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 05 04:41:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	759197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1341213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1126773	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	446837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	371216	57.321	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.640%
35) Dibromofluoromethane	7.634	113	330109	39.385	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.760%
50) Toluene-d8	10.103	98	1517303	48.597	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.200%
62) 4-Bromofluorobenzene	12.408	95	432028	42.788	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.580%
Target Compounds						
					Qvalue	
16) Acetone	3.879	43	2584	1.605	ug/l	98
20) Methylene Chloride	4.616	84	17254	2.220	ug/l	91
25) 2-Butanone	6.890	43	3616	1.943	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111325\
Data File : VY023779.D
Acq On : 13 Nov 2025 12:14
Operator : SY/MD
Sample : Q3615-01
Misc : 5.50g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
East Bathhouse- Concrete

Quant Time: Nov 13 13:32:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110425S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 05 04:41:50 2025
Response via : Initial Calibration

