

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY111419\
 Data File : VY000679.D
 Acq On : 14 Nov 2019 18:36
 Operator : SY/MD
 Sample : VY1114WBL01
 Misc : 5ML/MSVOA Y/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VBLK81

Quant Time: Nov 15 10:24:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\SOMYLM111419WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 15 10:16:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.69	114	337351	50.00	ug/L	0.00
28) Chlorobenzene-d5	11.49	117	310762	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.42	152	146269	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.24	65	117340	44.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.56%
7) Chloroethane-d5	2.76	69	104726	47.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.88%
11) 1,1-Dichloroethene-d2	3.85	63	165306	35.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.70%
21) 2-Butanone-d5	6.91	46	112674	107.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.79%
24) Chloroform-d	7.49	84	199961	52.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.72%
26) 1,2-Dichloroethane-d4	8.15	65	130060	48.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.54%
32) Benzene-d6	8.11	84	419751	48.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	9.13	67	135121	50.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.78%
41) Toluene-d8	10.18	98	382708	47.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.12%
43) trans-1,3-Dichloropropene-	10.44	79	65677	48.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.90%
47) 2-Hexanone-d5	10.79	63	78369	99.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.14%
57) 1,1,2,2-Tetrachloroethane-	12.56	84	145308	45.94	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.88%
64) 1,2-Dichlorobenzene-d4	13.72	152	138744	49.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.06%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY111419\
Data File : VY000679.D
Acq On : 14 Nov 2019 18:36
Operator : SY/MD
Sample : VY1114WBL01
Misc : 5ML/MSVOA Y/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VBLK81

Quant Time: Nov 15 10:24:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\SOMYLM111419WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 15 10:16:24 2019
Response via : Initial Calibration

