

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111422\
 Data File : VY011441.D
 Acq On : 14 Nov 2022 15:47
 Operator : KP/MD
 Sample : N5582-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 31878

Quant Time: Nov 15 01:50:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	211394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	365986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	329394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	144272	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	116009	65.489	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 130.980%	
35) Dibromofluoromethane	7.716	113	124210	60.559	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 121.120%	
50) Toluene-d8	10.179	98	434216	55.415	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.840%	
62) 4-Bromofluorobenzene	12.483	95	139831	51.163	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.320%	
Target Compounds						
20) Methylene Chloride	4.716	84	32250	6.117	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111422\
Data File : VY011441.D
Acq On : 14 Nov 2022 15:47
Operator : KP/MD
Sample : N5582-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
31878

Quant Time: Nov 15 01:50:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

