

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111422\
 Data File : VY011445.D
 Acq On : 14 Nov 2022 17:18
 Operator : KP/MD
 Sample : N5528-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SAMPLE-3

Quant Time: Nov 15 01:50:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 10 17:47:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	290989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	468109	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	328225	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	96017	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	156621	64.231	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.460%
35) Dibromofluoromethane	7.716	113	163753	62.420	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	124.840%
50) Toluene-d8	10.179	98	511969	51.084	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.160%
62) 4-Bromofluorobenzene	12.483	95	122353	35.002	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111422\
Data File : VY011445.D
Acq On : 14 Nov 2022 17:18
Operator : KP/MD
Sample : N5528-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SAMPLE-3

Quant Time: Nov 15 01:50:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111022S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 10 17:47:26 2022
Response via : Initial Calibration

