

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
 Data File : VY016337.D
 Acq On : 14 Nov 2023 10:02
 Operator : SY/MD
 Sample : VY1114SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1114SBL01

Quant Time: Nov 14 19:00:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	105738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.704	114	189310	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	176842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	80180	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	64766	67.018	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.040%
35) Dibromofluoromethane	7.734	113	58613	54.503	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.000%
50) Toluene-d8	10.191	98	220045	50.458	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.920%
62) 4-Bromofluorobenzene	12.496	95	75404	53.237	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
Data File : VY016337.D
Acq On : 14 Nov 2023 10:02
Operator : SY/MD
Sample : VY1114SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1114SBL01

Quant Time: Nov 14 19:00:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

