

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
 Data File : VY016340.D
 Acq On : 14 Nov 2023 11:47
 Operator : SY/MD
 Sample : 05346-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-110923

Quant Time: Nov 14 12:18:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	105422	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	185767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	179180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	81620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	67507	70.063	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.120%
35) Dibromofluoromethane	7.734	113	59466	56.351	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.700%
50) Toluene-d8	10.191	98	219652	51.328	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.660%
62) 4-Bromofluorobenzene	12.495	95	76183	54.813	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.620%
Target Compounds						
16) Acetone	3.954	43	1916	12.022	ug/l #	86
20) Methylene Chloride	4.729	84	6564	2.525	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
Data File : VY016340.D
Acq On : 14 Nov 2023 11:47
Operator : SY/MD
Sample : 05346-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-110923

Quant Time: Nov 14 12:18:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

