

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
 Data File : VY016341.D
 Acq On : 14 Nov 2023 12:11
 Operator : SY/MD
 Sample : 05385-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-11102023

Quant Time: Nov 14 19:02:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	10409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	15989	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	12893	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	4851	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	4454	46.818	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.640%
35) Dibromofluoromethane	7.728	113	4183	46.054	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.100%
50) Toluene-d8	10.191	98	18142	49.255	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.520%
62) 4-Bromofluorobenzene	12.489	95	4741	39.632	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
Data File : VY016341.D
Acq On : 14 Nov 2023 12:11
Operator : SY/MD
Sample : 05385-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-11102023

Quant Time: Nov 14 19:02:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

