

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
 Data File : VY016346.D
 Acq On : 14 Nov 2023 14:08
 Operator : SY/MD
 Sample : 05388-03
 Misc : 5.87g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3-VOC

Quant Time: Nov 15 05:28:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	96493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	172516	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	167822	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	77857	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	61722	69.987	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.980%	
35) Dibromofluoromethane	7.728	113	54562	55.676	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.360%	
50) Toluene-d8	10.191	98	203791	51.280	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.560%	
62) 4-Bromofluorobenzene	12.489	95	69953	54.196	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 108.400%	
Target Compounds						
					Qvalue	
16) Acetone	3.960	43	2127	14.581	ug/l #	83
25) 2-Butanone	7.015	43	1327	5.193	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
Data File : VY016346.D
Acq On : 14 Nov 2023 14:08
Operator : SY/MD
Sample : 05388-03
Misc : 5.87g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3-VOC

Quant Time: Nov 15 05:28:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

