

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
 Data File : VY016347.D
 Acq On : 14 Nov 2023 14:31
 Operator : SY/MD
 Sample : 05388-07
 Misc : 5.28g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-4-VOC

Quant Time: Nov 15 05:28:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	95517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	171056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	166021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	76369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	61083	69.970	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 139.940%		
35) Dibromofluoromethane	7.728	113	53958	55.529	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 111.060%		
50) Toluene-d8	10.191	98	202530	51.397	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 102.800%		
62) 4-Bromofluorobenzene	12.489	95	69687	54.451	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 108.900%		
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.722	84	5921	2.497	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
Data File : VY016347.D
Acq On : 14 Nov 2023 14:31
Operator : SY/MD
Sample : 05388-07
Misc : 5.28g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-4-VOC

Quant Time: Nov 15 05:28:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

