

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
 Data File : VY016348.D
 Acq On : 14 Nov 2023 14:55
 Operator : SY/MD
 Sample : 05385-01RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-11102023RE

Quant Time: Nov 15 05:28:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	10389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.704	114	16629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	15139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	6004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	5586	58.830	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.660%
35) Dibromofluoromethane	7.728	113	5335	56.477	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.960%
50) Toluene-d8	10.191	98	18851	49.211	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.420%
62) 4-Bromofluorobenzene	12.489	95	5380	43.242	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111423\
Data File : VY016348.D
Acq On : 14 Nov 2023 14:55
Operator : SY/MD
Sample : 05385-01RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-11102023RE

Quant Time: Nov 15 05:28:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

