

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy111519\
 Data File : VY000689.D
 Acq On : 15 Nov 2019 10:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 16 02:37:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y103019S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 30 13:30:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.492	0.522	-6.1	91	0.00
3 P	Chloromethane	0.678	0.683	-0.7	90	0.00
4 C	Vinyl Chloride	0.630	0.657	-4.3#	92	0.00
5 T	Bromomethane	0.378	0.539	-42.6#	134	0.00
6 T	Chloroethane	0.394	0.418	-6.1	96	0.00
7 T	Trichlorofluoromethane	0.850	0.939	-10.5	100	0.00
8 T	Diethyl Ether	0.302	0.316	-4.6	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.514	0.551	-7.2	98	0.00
10 T	Methyl Iodide	0.606	0.367	39.4#	51	0.00
11 T	Tert butyl alcohol	0.050	0.044	12.0	84	0.00
12 CM	1,1-Dichloroethene	0.499	0.538	-7.8#	93	0.00
13 T	Acrolein	0.044	0.033	25.0#	68	0.00
14 T	Allyl chloride	0.812	0.900	-10.8	98	0.00
15 T	Acrylonitrile	0.155	0.154	0.6	90	0.00
16 T	Acetone	0.143	0.204	-42.7#	117	0.00
17 T	Carbon Disulfide	1.575	1.618	-2.7	92	0.00
18 T	Methyl Acetate	0.443	0.443	0.0	96	0.00
19 T	Methyl tert-butyl Ether	1.389	1.444	-4.0	93	0.00
20 T	Methylene Chloride	0.611	0.602	1.5	96	0.00
21 T	trans-1,2-Dichloroethene	0.565	0.608	-7.6	96	0.00
22 T	Diisopropyl ether	1.765	1.946	-10.3	98	0.00
23 T	Vinyl Acetate	1.114	1.196	-7.4	95	0.00
24 P	1,1-Dichloroethane	0.956	1.034	-8.2	98	0.00
25 T	2-Butanone	0.228	0.260	-14.0	99	0.00
26 T	2,2-Dichloropropane	0.874	0.965	-10.4	102	0.00
27 T	cis-1,2-Dichloroethene	0.617	0.688	-11.5	100	0.00
28 T	Bromochloromethane	0.388	0.459	-18.3	88	0.00
29 T	Tetrahydrofuran	0.139	0.141	-1.4	92	0.00
30 C	Chloroform	0.967	1.085	-12.2#	101	0.00
31 T	Cyclohexane	1.002	1.034	-3.2	95	0.00
32 T	1,1,1-Trichloroethane	0.862	0.967	-12.2	100	0.00
33 S	1,2-Dichloroethane-d4	0.517	0.548	-6.0	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.294	0.302	-2.7	89	0.00
36 T	1,1-Dichloropropene	0.473	0.523	-10.6	100	0.00
37 T	Ethyl Acetate	0.281	0.284	-1.1	94	0.00
38 T	Carbon Tetrachloride	0.455	0.509	-11.9	99	0.00
39 T	Methylcyclohexane	0.611	0.647	-5.9	93	0.00
40 TM	Benzene	1.386	1.495	-7.9	97	0.00
41 T	Methacrylonitrile	0.157	0.169	-7.6	104	-0.01
42 TM	1,2-Dichloroethane	0.393	0.437	-11.2	99	0.00
43 T	Isopropyl Acetate	0.527	0.545	-3.4	94	0.00
44 TM	Trichloroethene	0.365	0.391	-7.1	97	0.00
45 C	1,2-Dichloropropane	0.341	0.382	-12.0#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy111519\
 Data File : VY000689.D
 Acq On : 15 Nov 2019 10:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 16 02:37:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y103019S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 30 13:30:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.188	0.200	-6.4	95	0.00
47 T	Bromodichloromethane	0.445	0.497	-11.7	97	0.00
48 T	Methyl methacrylate	0.230	0.244	-6.1	96	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	90	0.00
50 S	Toluene-d8	1.114	1.225	-10.0	88	0.00
51 T	4-Methyl-2-Pentanone	0.280	0.286	-2.1	90	0.00
52 CM	Toluene	0.876	0.975	-11.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.476	0.530	-11.3	95	0.00
54 T	cis-1,3-Dichloropropene	0.545	0.608	-11.6	98	0.00
55 T	1,1,2-Trichloroethane	0.271	0.287	-5.9	95	0.00
56 T	Ethyl methacrylate	0.385	0.414	-7.5	92	0.00
57 T	1,3-Dichloropropane	0.475	0.515	-8.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.174	0.134	23.0#	65	0.00
59 T	2-Hexanone	0.204	0.222	-8.8	93	0.00
60 T	Dibromochloromethane	0.311	0.345	-10.9	97	0.00
61 T	1,2-Dibromoethane	0.261	0.273	-4.6	91	0.00
62 S	4-Bromofluorobenzene	0.424	0.438	-3.3	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.413	0.425	-2.9	93	0.00
65 PM	Chlorobenzene	1.014	1.099	-8.4	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.352	0.390	-10.8	98	0.00
67 C	Ethyl Benzene	1.866	2.079	-11.4#	99	0.00
68 T	m/p-Xylenes	0.722	0.786	-8.9	96	0.00
69 T	o-Xylene	0.679	0.745	-9.7	99	0.00
70 T	Styrene	1.162	1.292	-11.2	98	0.00
71 P	Bromoform	0.202	0.218	-7.9	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.695	4.131	-11.8	99	0.00
74 T	N-amyl acetate	1.058	1.088	-2.8	88	0.00
75 P	1,1,2,2-Tetrachloroethane	0.732	0.801	-9.4	98	0.00
76 T	1,2,3-Trichloropropane	0.569	0.579	-1.8	81	0.00
77 T	Bromobenzene	0.838	0.911	-8.7	97	0.00
78 T	n-propylbenzene	4.459	5.103	-14.4	100	0.00
79 T	2-Chlorotoluene	2.475	2.765	-11.7	99	0.00
80 T	1,3,5-Trimethylbenzene	3.078	3.468	-12.7	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.254	0.279	-9.8	94	0.00
82 T	4-Chlorotoluene	2.566	2.926	-14.0	100	0.00
83 T	tert-Butylbenzene	2.628	3.002	-14.2	99	0.00
84 T	1,2,4-Trimethylbenzene	3.064	3.530	-15.2	98	0.00
85 T	sec-Butylbenzene	3.744	4.338	-15.9	99	0.00
86 T	p-Isopropyltoluene	3.337	3.804	-14.0	98	0.00
87 T	1,3-Dichlorobenzene	1.641	1.814	-10.5	97	0.00
88 T	1,4-Dichlorobenzene	1.640	1.800	-9.8	96	0.00
89 T	n-Butylbenzene	3.291	3.833	-16.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\vy111519\
Data File : VY000689.D
Acq On : 15 Nov 2019 10:19
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Nov 16 02:37:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y103019S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 30 13:30:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.617	0.710	-15.1	98	0.00
91 T	1,2-Dichlorobenzene	1.489	1.668	-12.0	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.129	0.131	-1.6	91	0.00
93 T	1,2,4-Trichlorobenzene	1.014	1.101	-8.6	98	0.00
94 T	Hexachlorobutadiene	0.523	0.574	-9.8	95	0.00
95 T	Naphthalene	2.251	2.357	-4.7	93	0.00
96 T	1,2,3-Trichlorobenzene	0.910	0.973	-6.9	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6