

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111521\
 Data File : VY006741.D
 Acq On : 15 Nov 2021 19:22
 Operator : SY/MD
 Sample : M4640-01
 Misc : 4.60G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2-AS

Quant Time: Nov 16 04:22:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	114904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	198480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	182483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	75685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	86594	57.501	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.000%
35) Dibromofluoromethane	7.728	113	63236	51.869	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.740%
50) Toluene-d8	10.185	98	252976	50.144	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.280%
62) 4-Bromofluorobenzene	12.483	95	78803	46.334	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.660%
Target Compounds					Qvalue	
16) Acetone	3.954	43	8730	19.080	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111521\
Data File : VY006741.D
Acq On : 15 Nov 2021 19:22
Operator : SY/MD
Sample : M4640-01
Misc : 4.60G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2-AS

Quant Time: Nov 16 04:22:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

