

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111521\
 Data File : VY006746.D
 Acq On : 15 Nov 2021 21:20
 Operator : SY/MD
 Sample : M4562-02
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T-1A

Quant Time: Nov 16 04:23:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	34304	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	53957	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	44945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	14673	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	21059	46.840	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.680%
35) Dibromofluoromethane	7.728	113	16266	49.079	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.160%
50) Toluene-d8	10.185	98	68402	49.875	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.740%
62) 4-Bromofluorobenzene	12.489	95	15756	34.078	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	68.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111521\
Data File : VY006746.D
Acq On : 15 Nov 2021 21:20
Operator : SY/MD
Sample : M4562-02
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T-1A

Quant Time: Nov 16 04:23:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

