

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111523\
 Data File : VY016361.D
 Acq On : 15 Nov 2023 13:14
 Operator : SY/MD
 Sample : 05408-07
 Misc : 7.28g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 710-PLAINFIELD-AVE

Quant Time: Nov 16 00:03:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	97030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	173910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	167967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	76057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	62970	71.007	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.020%
35) Dibromofluoromethane	7.734	113	56102	56.788	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.580%
50) Toluene-d8	10.191	98	204868	51.138	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.280%
62) 4-Bromofluorobenzene	12.495	95	71306	54.802	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.600%
Target Compounds					Qvalue	
52) Toluene	10.258	92	3367	1.105	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111523\
Data File : VY016361.D
Acq On : 15 Nov 2023 13:14
Operator : SY/MD
Sample : 05408-07
Misc : 7.28g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
710-PLAINFIELD-AVE

Quant Time: Nov 16 00:03:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

