

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111523\
 Data File : VY016362.D
 Acq On : 15 Nov 2023 13:37
 Operator : SY/MD
 Sample : 05408-09
 Misc : 8.41g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 710-RHINE-AVE

Quant Time: Nov 16 00:03:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	94872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	168532	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	163304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	73896	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	61638	71.086	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 142.180%	
35) Dibromofluoromethane	7.734	113	54312	56.731	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.460%	
50) Toluene-d8	10.191	98	199966	51.507	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 103.020%	
62) 4-Bromofluorobenzene	12.495	95	67869	53.825	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 107.640%	
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	1929	13.449	ug/l #	88
52) Toluene	10.252	92	5945	2.014	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111523\
Data File : VY016362.D
Acq On : 15 Nov 2023 13:37
Operator : SY/MD
Sample : 05408-09
Misc : 8.41g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
710-RHINE-AVE

Quant Time: Nov 16 00:03:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

