

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111523\
 Data File : VY016369.D
 Acq On : 15 Nov 2023 16:21
 Operator : SY/MD
 Sample : 05402-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-DRUM-COMP

Quant Time: Nov 16 00:06:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 09 00:45:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	94449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.704	114	169432	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	162463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	74018	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	60803	70.437	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.880%
35) Dibromofluoromethane	7.728	113	54389	56.509	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.020%
50) Toluene-d8	10.191	98	199567	51.131	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.260%
62) 4-Bromofluorobenzene	12.489	95	69689	54.975	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.940%
Target Compounds						Qvalue
20) Methylene Chloride	4.735	84	9186	6.047	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111523\
Data File : VY016369.D
Acq On : 15 Nov 2023 16:21
Operator : SY/MD
Sample : 05402-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-DRUM-COMP

Quant Time: Nov 16 00:06:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110823S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 09 00:45:52 2023
Response via : Initial Calibration

