

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
 Data File : VY006766.D
 Acq On : 16 Nov 2021 17:01
 Operator : SY/MD
 Sample : M4654-02RE
 Misc : 6.14G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PARKINGLOT-BOTTOM-U-SP-1-1

Quant Time: Nov 17 04:08:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.801	168	92552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	162270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	152619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	66725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	73367	60.484	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 120.960%			
35) Dibromofluoromethane	7.728	113	50284	50.449	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 100.900%			
50) Toluene-d8	10.185	98	211275	51.223	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 102.440%			
62) 4-Bromofluorobenzene	12.483	95	72064	51.826	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 103.660%			
Target Compounds					Qvalue	
16) Acetone	3.954	43	3627	9.841	ug/l	97
20) Methylene Chloride	4.735	84	2561	Below Cal	#	80
25) 2-Butanone	6.997	43	384	12.184	ug/l #	52
31) Cyclohexane	7.801	56	3646	1.710	ug/l #	56
36) 1,1-Dichloropropene	7.801	75	7650	4.337	ug/l #	52
38) Carbon Tetrachloride	7.795	117	9225	5.113	ug/l #	16
39) Methylcyclohexane	9.191	83	14652	6.811	ug/l	95
51) 4-Methyl-2-Pentanone	9.965	43	18115	15.547	ug/l #	44
55) 1,1,2-Trichloroethane	10.624	97	19213	21.569	ug/l #	21
56) Ethyl methacrylate	10.526	69	3680	2.658	ug/l #	1
59) 2-Hexanone	10.807	43	25808	51.125	ug/l #	33
67) Ethyl Benzene	11.599	91	10449	1.775	ug/l	96
69) o-Xylene	12.032	106	13816	6.502	ug/l	92
73) Isopropylbenzene	12.331	105	9309	1.832	ug/l	96
74) N-amyl acetate	12.124	43	98319	53.189	ug/l #	62
75) 1,1,2,2-Tetrachloroethane	12.526	83	4680	4.978	ug/l #	1
76) 1,2,3-Trichloropropane	12.483	75	40849	56.309	ug/l #	100
78) n-propylbenzene	12.672	91	28347	4.576	ug/l	97
79) 2-Chlorotoluene	12.770	91	10450	3.083	ug/l #	45
80) 1,3,5-Trimethylbenzene	12.813	105	84105	20.195	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.483	75	40849	111.574	ug/l #	9
82) 4-Chlorotoluene	12.813	91	12958	3.657	ug/l #	62
85) sec-Butylbenzene	13.258	105	33900	6.293	ug/l	86
86) p-Isopropyltoluene	13.373	119	31450	6.897	ug/l	97
89) n-Butylbenzene	13.697	91	31609	7.346	ug/l #	69
90) Hexachloroethane	13.977	117	11682	14.215	ug/l #	6
92) 1,2-Dibromo-3-Chloropr...	14.337	75	782	4.706	ug/l #	5
95) Naphthalene	15.215	128	6219	2.587	ug/l #	1
96) 1,2,3-Trichlorobenzene	15.452	180	7839	7.240	ug/l #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
Data File : VY006766.D
Acq On : 16 Nov 2021 17:01
Operator : SY/MD
Sample : M4654-02RE
Misc : 6.14G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1

Quant Time: Nov 17 04:08:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.801 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion:168 Resp: 92552
Ion Ratio Lower Upper
168 100
99 54.1 44.8 67.2



#16
Acetone
Concen: 9.841 ug/l
RT: 3.954 min Scan# 350
Delta R.T. -0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01


Tgt Ion: 43 Resp: 3627
Ion Ratio Lower Upper
43 100
58 26.9 22.9 34.3








#20
Methylene Chloride
Concen: Below Cal
RT: 4.735 min Scan# 41
Delta R.T. 0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1







Tgt Ion: 84 Resp: 2561
Ion Ratio Lower Upper
84 100
49 110.4 107.8 161.8
51 31.5 32.8 49.2#
86 45.4 50.6 75.8#



#25
2-Butanone
Concen: 12.184 ug/l
RT: 6.997 min Scan# 849
Delta R.T. 0.000 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 43 Resp: 384
Ion Ratio Lower Upper
43 100
72 0.0 18.9 28.3#









#31
Cyclohexane
Concen: 1.710 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.000 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 56 Resp: 3646
Ion Ratio Lower Upper
56 100
69 109.7 26.6 40.0#
84 86.8 65.0 97.4



#33
1,2-Dichloroethane-d4
Concen: 60.484 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. -0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 65 Resp: 73367
Ion Ratio Lower Upper
65 100
67 50.2 0.0 98.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.697 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VY006766.D

Acq: 16 Nov 2021 17:01

Instrument :

MSVOA_Y

ClientSampleId :

PARKINGLOT-BOTTOM-U-SP-1-1





Tgt Ion:114 Resp: 162270

Ion Ratio Lower Upper

114 100

63 21.2 0.0 45.0

88 15.4 0.0 30.6



#35

Dibromofluoromethane

Concen: 50.449 ug/l

RT: 7.728 min Scan# 969

Delta R.T. 0.000 min

Lab File: VY006766.D

Acq: 16 Nov 2021 17:01




Tgt Ion:113 Resp: 50284


Ion Ratio Lower Upper

113 100

111 101.6 82.3 123.5

192 21.6 17.4 26.2





#36

1,1-Dichloropropene

Concen: 4.337 ug/l

RT: 7.801 min Scan# 91

Delta R.T. -0.128 min

Lab File: VY006766.D

Acq: 16 Nov 2021 17:01





Instrument :

MSVOA_Y

ClientSampleId :

PARKINGLOT-BOTTOM-U-SP-1-1

Tgt Ion:	75	Resp:	7650
Ion	Ratio	Lower	Upper
75	100		
110	10.9	17.1	51.3#
77	0.0	24.9	37.3#



#38

Carbon Tetrachloride

Concen: 5.113 ug/l




RT: 7.795 min Scan# 980


Delta R.T. -0.116 min

Lab File: VY006766.D

Acq: 16 Nov 2021 17:01

Tgt Ion:	117	Resp:	9225
Ion	Ratio	Lower	Upper
117	100		
119	4.7	76.6	115.0#
121	0.0	24.7	37.1#





#39

Methylcyclohexane

Concen: 6.811 ug/l

RT: 9.191 min Scan# 1209

Delta R.T. -0.006 min

Lab File: VY006766.D

Acq: 16 Nov 2021 17:01

Instrument :

MSVOA_Y

ClientSampleId :

PARKINGLOT-BOTTOM-U-SP-1-1





Tgt Ion: 83 Resp: 14652

Ion Ratio Lower Upper

83 100

55 78.5 65.3 97.9

98 50.4 37.0 55.4



#50

Toluene-d8

Concen: 51.223 ug/l

RT: 10.185 min Scan# 1372

Delta R.T. 0.000 min

Lab File: VY006766.D




Acq: 16 Nov 2021 17:01


Tgt Ion: 98 Resp: 211275

Ion Ratio Lower Upper

98 100

100 62.8 50.9 76.3









#51
4-Methyl-2-Pentanone
Concen: 15.547 ug/l
RT: 9.965 min Scan# 11
Delta R.T. -0.110 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 43 Resp: 18115
Ion Ratio Lower Upper
43 100
58 2.9 28.7 43.1#



#55
1,1,2-Trichloroethane
Concen: 21.569 ug/l
RT: 10.624 min Scan# 1444
Delta R.T. -0.030 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 97 Resp: 19213
Ion Ratio Lower Upper
97 100
83 8.0 66.2 99.2#
85 2.6 42.2 63.4#
99 0.0 48.6 73.0#









#56
Ethyl methacrylate
Concen: 2.658 ug/l
RT: 10.526 min Scan# 1428
Delta R.T. 0.012 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 69 Resp: 3680
Ion Ratio Lower Upper
69 100
41 167.6 62.9 94.3#
39 111.8 35.7 53.5#



#59
2-Hexanone
Concen: 51.125 ug/l
RT: 10.807 min Scan# 1474
Delta R.T. -0.030 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 43 Resp: 25808
Ion Ratio Lower Upper
43 100
58 3.1 24.6 73.6#









#62
4-Bromofluorobenzene
Concen: 51.826 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 95 Resp: 72064
Ion Ratio Lower Upper
95 100
174 81.0 0.0 177.0
176 80.4 0.0 169.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 117 Resp: 152619
Ion Ratio Lower Upper
117 100
82 57.2 45.7 68.5
119 33.1 25.5 38.3









#67
Ethyl Benzene
Concen: 1.775 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. 0.000 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 91 Resp: 10449
Ion Ratio Lower Upper
91 100
106 32.4 24.2 36.4



#69
o-Xylene
Concen: 6.502 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. -0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 106 Resp: 13816
Ion Ratio Lower Upper
106 100
91 233.3 110.3 330.9









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument : MSVOA_Y
ClientSampleId : PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion:152 Resp: 66725
Ion Ratio Lower Upper
152 100
115 62.6 28.5 85.5
150 153.9 0.0 348.2



#73
Isopropylbenzene
Concen: 1.832 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. -0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion:105 Resp: 9309
Ion Ratio Lower Upper
105 100
120 24.6 13.3 39.9









#74
N-amy1 acetate
Concen: 53.189 ug/l
RT: 12.124 min Scan# 1
Delta R.T. -0.024 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion:	43	Resp:	98319
Ion	Ratio	Lower	Upper
43	100		
70	8.9	29.4	44.0#
55	14.7	18.1	27.1#
61	0.0	17.6	26.4#



#75
1,1,2,2-Tetrachloroethane
Concen: 4.978 ug/l
RT: 12.526 min Scan# 1756
Delta R.T. -0.061 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion:	83	Resp:	4680
Ion	Ratio	Lower	Upper
83	100		
131	0.0	5.5	16.4#
85	202.0	31.9	95.9#









#76
1,2,3-Trichloropropane
Concen: 56.309 ug/l
RT: 12.483 min Scan# 11
Delta R.T. -0.153 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 75 Resp: 40849
Ion Ratio Lower Upper
75 100
77 1.7 0.0 0.0#



#78
n-propylbenzene
Concen: 4.576 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. -0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 91 Resp: 28347
Ion Ratio Lower Upper
91 100
120 21.2 11.3 33.8









#79
2-Chlorotoluene
Concen: 3.083 ug/l
RT: 12.770 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 91 Resp: 10450
Ion Ratio Lower Upper
91 100
126 2.4 17.1 51.2#



#80
1,3,5-Trimethylbenzene
Concen: 20.195 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. -0.006 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 105 Resp: 84105
Ion Ratio Lower Upper
105 100
120 48.5 24.1 72.4









#81
trans-1,4-Dichloro-2-butene
Concen: 111.574 ug/l
RT: 12.483 min Scan# 11
Delta R.T. 0.098 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01




Instrument :
MSVOA_Y
ClientSampleId :
PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 75 Resp: 40849
Ion Ratio Lower Upper
75 100
53 2.0 85.8 128.6#
89 0.9 39.3 58.9#



#82
4-Chlorotoluene
Concen: 3.657 ug/l
RT: 12.813 min Scan# 1803
Delta R.T. -0.049 min
Lab File: VY006766.D
Acq: 16 Nov 2021 17:01

Tgt Ion: 91 Resp: 12958
Ion Ratio Lower Upper
91 100
126 11.9 16.8 50.4#









#85
 sec-Butylbenzene
 Concen: 6.293 ug/l
 RT: 13.258 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: VY006766.D
 Acq: 16 Nov 2021 17:01




Instrument :
 MSVOA_Y
 ClientSampleId :
 PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion:105 Resp: 33900
 Ion Ratio Lower Upper
 105 100
 134 26.6 10.1 30.3



#86
 p-Isopropyltoluene
 Concen: 6.897 ug/l
 RT: 13.373 min Scan# 1895
 Delta R.T. 0.000 min
 Lab File: VY006766.D
 Acq: 16 Nov 2021 17:01

Tgt Ion:119 Resp: 31450
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.1 39.1
 91 27.2 12.4 37.4





#89

n-Butylbenzene

Concen: 7.346 ug/l

RT: 13.697 min Scan# 1948

Delta R.T. -0.006 min

Lab File: VY006766.D

Acq: 16 Nov 2021 17:01

Instrument :

MSVOA_Y

ClientSampleId :

PARKINGLOT-BOTTOM-U-SP-1-1





Tgt Ion: 91 Resp: 31609

Ion Ratio Lower Upper

91 100

92 67.3 26.3 78.9

134 0.0 12.9 38.7#



#90

Hexachloroethane

Concen: 14.215 ug/l

RT: 13.977 min Scan# 1994

Delta R.T. 0.012 min

Lab File: VY006766.D




Acq: 16 Nov 2021 17:01


Tgt Ion: 117 Resp: 11682

Ion Ratio Lower Upper

117 100

201 0.0 43.4 130.1#









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 4.706 ug/l
 RT: 14.337 min Scan# 2053
 Delta R.T. -0.024 min
 Lab File: VY006766.D
 Acq: 16 Nov 2021 17:01




Instrument : MSVOA_Y
 ClientSampleId : PARKINGLOT-BOTTOM-U-SP-1-1


Tgt Ion: 75 Resp: 782
 Ion Ratio Lower Upper
 75 100
 155 0.0 39.0 116.9#
 157 0.0 49.2 147.6#



#95
 Naphthalene
 Concen: 2.587 ug/l
 RT: 15.215 min Scan# 2197
 Delta R.T. -0.024 min
 Lab File: VY006766.D
 Acq: 16 Nov 2021 17:01

Tgt Ion: 128 Resp: 6219
 Ion Ratio Lower Upper
 128 100
 127 133.8 11.0 16.4#
 129 63.7 8.9 13.3#





#96
 1,2,3-Trichlorobenzene
 Concen: 7.240 ug/l
 RT: 15.452 min Scan# 21
 Delta R.T. 0.024 min
 Lab File: VY006766.D
 Acq: 16 Nov 2021 17:01

Instrument :
 MSVOA_Y
 ClientSampleId :
 PARKINGLOT-BOTTOM-U-SP-1-1

Tgt Ion	Ratio	Lower	Upper
180	100		
182	0.0	47.8	143.4#
145	422.1	15.3	45.8#

