

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
 Data File : VY006754.D
 Acq On : 16 Nov 2021 12:17
 Operator : SY/MD
 Sample : M4687-02
 Misc : 6.08G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VWE-B35-111221-0-1

Quant Time: Nov 17 04:02:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	99714	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	172876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	153520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	54097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	79897	61.136	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.280%
35) Dibromofluoromethane	7.728	113	41768	39.334	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.660%
50) Toluene-d8	10.185	98	217763	49.557	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.120%
62) 4-Bromofluorobenzene	12.483	95	62785	42.383	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.760%
Target Compounds						
						Qvalue
16) Acetone	3.960	43	18654	46.980	ug/l	100
25) 2-Butanone	6.996	43	4333	18.660	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
Data File : VY006754.D
Acq On : 16 Nov 2021 12:17
Operator : SY/MD
Sample : M4687-02
Misc : 6.08G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VWE-B35-111221-0-1

Quant Time: Nov 17 04:02:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 16 04:20:03 2021
Response via : Initial Calibration

