

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
 Data File : VY006755.D
 Acq On : 16 Nov 2021 12:41
 Operator : SY/MD
 Sample : M4687-03
 Misc : 5.29G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampleId :
 VWE-B35-111221-9-10

Quant Time: Nov 17 04:02:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	99567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	172341	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	157090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	81262	62.272	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.540%
35) Dibromofluoromethane	7.728	113	40127	37.906	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.820%
50) Toluene-d8	10.185	98	216265	49.369	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.489	95	67271	45.552	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.100%
Target Compounds						
25) 2-Butanone	6.996	43	920	13.023	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111621\
 Data File : VY006755.D
 Acq On : 16 Nov 2021 12:41
 Operator : SY/MD
 Sample : M4687-03
 Misc : 5.29G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VWE-B35-111221-9-10

Quant Time: Nov 17 04:02:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y111521S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 16 04:20:03 2021
 Response via : Initial Calibration

